



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041025\
Data File : VX045685.D
Acq On : 10 Apr 2025 08:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	24.6	7798	PASS
75	95	30	60	58.1	18399	PASS
95	95	100	100	100.0	31677	PASS
96	95	5	9	6.6	2088	PASS
173	174	0.00	2	1.0	210	PASS
174	95	50	100	69.4	21989	PASS
175	174	5	9	7.7	1702	PASS
176	174	95	101	97.7	21480	PASS
177	176	5	9	6.4	1384	PASS