

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041119\
 Data File : VX008847.D
 Acq On : 11 Apr 2019 13:25
 Operator : JC/SP
 Sample : VX0411WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0411WBL01

Quant Time: Apr 12 07:49:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	176252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	283455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	239072	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	97351	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115201	44.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.62%	
35) Dibromofluoromethane	5.49	113	84359	46.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.10%	
50) Toluene-d8	8.71	98	349688	48.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.13	95	113164	44.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.44%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041119\
Data File : VX008847.D
Acq On : 11 Apr 2019 13:25
Operator : JC/SP
Sample : VX0411WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0411WBL01

Quant Time: Apr 12 07:49:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

