

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041119\
 Data File : VX008856.D
 Acq On : 11 Apr 2019 16:54
 Operator : JC/SP
 Sample : K2349-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW206-190408

Quant Time: Apr 12 08:18:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	197242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	318894	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	274990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	129267	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
35) Dibromofluoromethane	5.49	113	93976	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	8.71	98	399269	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	131702	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
16) Acetone	2.44	43	2017	1.217	ug/l	93
44) Trichloroethene	7.21	130	9587	3.964	ug/l	95
64) Tetrachloroethene	9.34	164	21887	10.743	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041119\
Data File : VX008856.D
Acq On : 11 Apr 2019 16:54
Operator : JC/SP
Sample : K2349-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW206-190408

Quant Time: Apr 12 08:18:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

