

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041119\
 Data File : VX008862.D
 Acq On : 11 Apr 2019 19:14
 Operator : JC/SP
 Sample : K2349-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW071-190409

Quant Time: Apr 12 08:31:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	193071	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	309703	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	270466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116906	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126256	44.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.66%	
35) Dibromofluoromethane	5.50	113	92620	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
50) Toluene-d8	8.71	98	391620	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	11.13	95	132552	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	29253	10.538	ug/l	91
44) Trichloroethene	7.21	130	100897	42.954	ug/l	98
64) Tetrachloroethene	9.34	164	770032	384.272	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041119\
Data File : VX008862.D
Acq On : 11 Apr 2019 19:14
Operator : JC/SP
Sample : K2349-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW071-190409

Quant Time: Apr 12 08:31:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

