

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041119\
 Data File : VX008867.D
 Acq On : 11 Apr 2019 21:10
 Operator : JC/SP
 Sample : K2347-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BCOG01

Quant Time: Apr 12 08:44:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	195185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	317192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	281380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	123317	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	131196	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
35) Dibromofluoromethane	5.50	113	94472	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
50) Toluene-d8	8.71	98	402920	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
62) 4-Bromofluorobenzene	11.13	95	137676	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
Target Compounds						
16) Acetone	2.44	43	37787	23.045	ug/l	99
20) Methylene Chloride	2.86	84	6470	2.357	ug/l	97
43) Isopropyl Acetate	6.45	43	9768	1.347	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041119\
Data File : VX008867.D
Acq On : 11 Apr 2019 21:10
Operator : JC/SP
Sample : K2347-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BCOG01

Quant Time: Apr 12 08:44:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

