

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041124\
 Data File : VX040966.D
 Acq On : 11 Apr 2024 10:00
 Operator : JC/MD
 Sample : VX0411WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK667

Quant Time: Apr 12 01:45:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 11 05:36:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	302240	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	295054	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	169248	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	99276	46.181	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.360%
7) Chloroethane-d5	1.666	69	85792	48.933	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.860%
11) 1,1-Dichloroethene-d2	2.307	65	41466	45.179	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.360%
21) 2-Butanone-d5	4.446	46	131802	112.554	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.550%
24) Chloroform-d	5.056	84	180418	44.979	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.960%
26) 1,2-Dichloroethane-d4	5.952	65	119631	45.851	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.700%
32) Benzene-d6	5.971	84	361814	45.169	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.340%
36) 1,2-Dichloropropane-d6	7.306	67	105971	45.361	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.720%
41) Toluene-d8	8.647	98	350724	44.817	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.640%
43) trans-1,3-Dichloroprop...	8.946	79	52205	44.405	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.820%
47) 2-Hexanone-d5	9.385	63	92674	94.088	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.090%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	160443	46.362	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.720%
66) 1,2-Dichlorobenzene-d4	12.317	152	158730	46.753	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.500%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041124\
Data File : VX040966.D
Acq On : 11 Apr 2024 10:00
Operator : JC/MD
Sample : VX0411WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK667

Quant Time: Apr 12 01:45:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 11 05:36:34 2024
Response via : Initial Calibration

