

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041124\
 Data File : VX040989.D
 Acq On : 11 Apr 2024 20:53
 Operator : JC/MD
 Sample : VX0411WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 VBLK668

Quant Time: Apr 12 01:48:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 11 05:36:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	255519	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	229895	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	107391	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	69952	38.490	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	76.980%		
7) Chloroethane-d5	1.648	69	35504	23.953	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery =	47.900%#		
11) 1,1-Dichloroethene-d2	2.294	65	31501	40.598	ug/L	-0.01
Spiked Amount	50.000	Range 60 - 125	Recovery =	81.200%		
21) 2-Butanone-d5	4.465	46	119279	120.485	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery =	120.480%		
24) Chloroform-d	5.056	84	148945	43.923	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	87.840%		
26) 1,2-Dichloroethane-d4	5.952	65	100114	45.386	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	90.780%		
32) Benzene-d6	5.971	84	309453	49.582	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	99.160%		
36) 1,2-Dichloropropane-d6	7.306	67	92565	50.853	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	101.700%		
41) Toluene-d8	8.647	98	280533	46.008	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	92.020%		
43) trans-1,3-Dichloroprop...	8.952	79	37632	41.082	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	82.160%		
47) 2-Hexanone-d5	9.391	63	86426	112.615	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery =	112.610%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	140045	51.938	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	103.880%		
66) 1,2-Dichlorobenzene-d4	12.317	152	116294	53.983	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	107.960%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041124\
 Data File : VX040989.D
 Acq On : 11 Apr 2024 20:53
 Operator : JC/MD
 Sample : VX0411WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK668

Quant Time: Apr 12 01:48:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 11 05:36:34 2024
 Response via : Initial Calibration

