



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041124\
Data File : VX040964.D
Acq On : 11 Apr 2024 08:29
Operator : JC/MD
Sample : BFB670
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB670

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM041024WMA.M
Title : VOC Analysis
Last Update : Thu Apr 11 05:36:34 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.6	9876	PASS
75	95	30	80	50.2	26659	PASS
95	95	100	100	100.0	53059	PASS
96	95	5	9	6.8	3584	PASS
173	174	0.00	2	1.1	577	PASS
174	95	50	120	100.6	53352	PASS
175	174	5	9	7.6	4040	PASS
176	174	95	101	98.2	52400	PASS
177	176	5	9	6.2	3245	PASS

m/z	Abundance
39.90	165.0
44.00	121.0

Instrument :
MSVOA_X
ClientSampleId :
BFB670

m/z	Abundance
35.90	434.0
37.10	2567.0
38.00	2407.0
39.10	758.0
40.00	254.0
44.10	343.0
45.00	439.0
47.10	606.0
47.90	314.0
49.00	2251.0
50.00	10268.0
51.10	3198.0
52.00	187.0
54.90	101.0
56.00	775.0
57.00	1497.0
58.00	91.0
59.90	542.0
61.00	2597.0
62.00	2560.0
63.00	1997.0
64.00	308.0
65.10	117.0
66.90	210.0
68.00	5543.0
69.00	5481.0
70.00	307.0
72.00	396.0
73.00	2370.0
74.00	9346.0
75.00	25856.0
76.00	2583.0
77.00	243.0
77.80	158.0
78.90	1623.0
79.90	497.0
80.90	1759.0
81.90	444.0
86.00	87.0
86.90	1861.0
88.00	2139.0
90.90	233.0
92.00	1451.0
92.90	2341.0
94.00	6068.0
95.00	50264.0
96.00	3724.0
97.00	108.0
103.80	309.0
105.00	69.0
105.90	191.0
107.00	62.0
114.90	70.0
115.90	260.0
116.80	371.0
117.80	193.0
118.70	222.0
127.80	241.0
129.00	111.0
129.80	144.0
134.90	71.0
136.90	98.0
140.90	534.0
142.10	114.0
142.80	577.0
144.90	96.0
145.90	102.0
147.90	85.0
148.80	57.0
149.80	77.0
152.90	57.0
154.90	152.0
157.00	111.0
158.80	56.0
172.00	308.0
173.90	43896.0
174.90	3203.0
175.90	42512.0
176.90	2546.0
177.90	140.0

Instrument :
MSVOA_X
ClientSampleId :
BFB670

m/z	Abundance
36.00	533.0
37.00	2966.0
38.00	2934.0
39.00	1041.0
40.10	221.0
44.10	350.0
45.00	572.0
47.00	669.0
48.00	471.0
49.00	2361.0
50.00	11687.0
51.00	3594.0
51.90	140.0
55.00	163.0
56.00	1035.0
57.00	1658.0
58.10	89.0
60.00	500.0
61.00	3054.0
62.00	3095.0
63.00	2582.0
64.10	354.0
64.90	226.0
67.10	204.0
68.00	6357.0
69.00	6438.0
70.00	530.0
72.00	441.0
73.00	2783.0
74.00	11327.0
75.00	32088.0
76.00	2732.0
77.00	375.0
77.90	262.0
78.90	2112.0
79.90	606.0
80.90	2043.0
81.90	442.0
86.00	135.0
87.00	2517.0
88.00	2561.0
90.90	323.0
92.00	1877.0
93.00	3150.0
94.00	8237.0
95.00	63784.0
96.00	4280.0
97.00	172.0
103.80	297.0
104.90	113.0
105.70	265.0
107.00	73.0
114.80	116.0
115.90	271.0
116.90	522.0
117.80	346.0
118.90	344.0
127.90	191.0
129.00	122.0
129.90	262.0
130.90	73.0
134.80	119.0
136.90	155.0
140.80	837.0
142.00	86.0
142.90	947.0
144.80	112.0
145.70	141.0
147.90	159.0
148.80	84.0
149.80	123.0
152.80	51.0
154.70	194.0
156.90	185.0
158.80	88.0
160.80	79.0
171.90	417.0
173.00	957.0
173.90	63248.0
175.00	5040.0
175.90	61176.0
177.00	3900.0
177.90	151.0

Instrument :
MSVOA_X
ClientSampleId :
BFB670

m/z	Abundance
36.00	370.0
37.00	1939.0
38.00	1674.0
39.00	616.0
39.90	144.0
44.00	379.0
45.00	315.0
47.00	493.0
48.00	242.0
49.00	1673.0
50.00	7674.0
51.00	2356.0
52.00	109.0
54.90	86.0
56.00	797.0
57.00	1189.0
58.10	51.0
60.00	607.0
61.00	2134.0
62.00	1902.0
63.00	1631.0
63.90	179.0
65.00	154.0
67.00	163.0
68.00	4506.0
69.00	4524.0
70.00	377.0
72.00	240.0
73.00	2033.0
74.00	8421.0
75.00	22032.0
76.00	1798.0
77.00	288.0
77.90	204.0
78.90	1389.0
79.90	447.0
80.90	1543.0
81.90	337.0
86.10	91.0
87.00	1729.0
88.00	1853.0
91.00	179.0
92.00	1228.0
93.00	2025.0
94.00	5598.0
95.00	45128.0
96.00	2747.0
97.00	103.0
103.80	245.0
104.90	110.0
105.90	284.0
106.80	63.0
115.90	293.0
116.80	290.0
117.80	121.0
118.90	253.0
127.80	188.0
129.00	72.0
129.90	196.0
130.90	74.0
134.90	84.0
136.70	90.0
140.90	552.0
142.00	52.0
142.90	715.0
144.80	109.0
145.70	108.0
147.80	122.0
149.80	64.0
154.90	111.0
156.80	102.0
158.90	67.0
160.90	59.0
171.20	113.0
172.00	349.0
173.00	775.0
173.90	52912.0
175.00	3878.0
175.90	53512.0
176.90	3289.0
178.00	70.0

Instrument :
MSVOA_X
ClientSampleId :
BFB670