

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
 Data File : VX045731.D
 Acq On : 11 Apr 2025 11:09
 Operator : JC/MD
 Sample : VX0411MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0411MBL01

Quant Time: Apr 11 23:03:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123400	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63273	54.115	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.220%	
35) Dibromofluoromethane	5.379	113	45159	51.579	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.160%	
50) Toluene-d8	8.647	98	156210	51.117	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.240%	
62) 4-Bromofluorobenzene	11.079	95	61301	55.071	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 110.140%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
Data File : VX045731.D
Acq On : 11 Apr 2025 11:09
Operator : JC/MD
Sample : VX0411MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0411MBL01

Quant Time: Apr 11 23:03:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

