

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041125\  
 Data File : VX045732.D  
 Acq On : 11 Apr 2025 11:32  
 Operator : JC/MD  
 Sample : VX0411WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0411WBL01

Quant Time: Apr 11 23:03:38 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 02 03:11:43 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 65554    | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 128675   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.049 | 117   | 118153   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 48353    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 64154    | 53.515   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | = 107.020% |          |
| 35) Dibromofluoromethane    | 5.379  | 113   | 46741    | 51.198   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 102.400% |          |
| 50) Toluene-d8              | 8.647  | 98    | 159396   | 50.021   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | = 100.040% |          |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 58527    | 50.424   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | = 100.840% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041125\  
Data File : VX045732.D  
Acq On : 11 Apr 2025 11:32  
Operator : JC/MD  
Sample : VX0411WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0411WBL01

Quant Time: Apr 11 23:03:38 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 02 03:11:43 2025  
Response via : Initial Calibration

