

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
 Data File : VX045735.D
 Acq On : 11 Apr 2025 12:45
 Operator : JC/MD
 Sample : Q1773-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT190D2-GW-20250407

Quant Time: Apr 11 23:04:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	112481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	44175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62219	52.734	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.460%	
35) Dibromofluoromethane	5.379	113	44969	50.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.980%	
50) Toluene-d8	8.647	98	155879	50.638	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.280%	
62) 4-Bromofluorobenzene	11.079	95	55472	49.472	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.940%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	15502	19.557	ug/l	100
12) 1,1-Dichloroethene	2.312	96	2976	3.842	ug/l	94
16) Acetone	2.379	43	575	1.187	ug/l #	88
27) cis-1,2-Dichloroethene	4.489	96	1620	1.680	ug/l	99
44) Trichloroethene	7.123	130	112715	130.395	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
Data File : VX045735.D
Acq On : 11 Apr 2025 12:45
Operator : JC/MD
Sample : Q1773-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT190D2-GW-20250407

Quant Time: Apr 11 23:04:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

