

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
 Data File : VX045736.D
 Acq On : 11 Apr 2025 13:08
 Operator : JC/MD
 Sample : Q1773-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT190D1-GW-20250409

Quant Time: Apr 11 23:04:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66855	54.454	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.900%	
35) Dibromofluoromethane	5.385	113	47739	51.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.680%	
50) Toluene-d8	8.647	98	164412	50.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.320%	
62) 4-Bromofluorobenzene	11.079	95	61082	51.670	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.340%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	7245	8.784	ug/l	95
12) 1,1-Dichloroethene	2.318	96	2974	3.690	ug/l	89
27) cis-1,2-Dichloroethene	4.489	96	2777	2.768	ug/l	89
30) Chloroform	5.092	83	2443	1.394	ug/l	97
38) Carbon Tetrachloride	5.678	117	4109	3.028	ug/l #	80
44) Trichloroethene	7.122	130	404492	443.842	ug/l	94
55) 1,1,2-Trichloroethane	9.159	97	2023	2.188	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
Data File : VX045736.D
Acq On : 11 Apr 2025 13:08
Operator : JC/MD
Sample : Q1773-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT190D1-GW-20250409

Quant Time: Apr 11 23:04:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

