

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
 Data File : VX045741.D
 Acq On : 11 Apr 2025 15:04
 Operator : JC/MD
 Sample : Q1782-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Apr 11 23:05:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132692	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66987	53.302	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.600%	
35) Dibromofluoromethane	5.379	113	48989	52.036	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.080%	
50) Toluene-d8	8.647	98	167371	50.934	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.860%	
62) 4-Bromofluorobenzene	11.079	95	60578	50.611	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	101.220%	
Target Compounds						
16) Acetone	2.379	43	915	1.773	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
Data File : VX045741.D
Acq On : 11 Apr 2025 15:04
Operator : JC/MD
Sample : Q1782-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Apr 11 23:05:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

