

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
 Data File : VX045742.D
 Acq On : 11 Apr 2025 15:28
 Operator : JC/MD
 Sample : Q1782-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Apr 11 23:05:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	65299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128436	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66071	55.329	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.660%	
35) Dibromofluoromethane	5.385	113	46646	51.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.380%	
50) Toluene-d8	8.646	98	163762	51.487	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.980%	
62) 4-Bromofluorobenzene	11.079	95	59714	51.542	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.080%	
Target Compounds						
16) Acetone	2.385	43	724	1.476	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
Data File : VX045742.D
Acq On : 11 Apr 2025 15:28
Operator : JC/MD
Sample : Q1782-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Apr 11 23:05:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

