

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
 Data File : VX045749.D
 Acq On : 11 Apr 2025 18:10
 Operator : JC/MD
 Sample : Q1790-04 20X
 Misc : 6.84g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-871-G1-SO-23.0-23.5-041025

Quant Time: Apr 12 00:21:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122428	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61995	54.200	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.400%
35) Dibromofluoromethane	5.385	113	43714	50.325	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.660%
50) Toluene-d8	8.647	98	151283	49.898	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.079	95	57868	52.400	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.800%
Target Compounds						
20) Methylene Chloride	2.788	84	247	0.281	ug/l #	66
44) Trichloroethene	7.129	130	2315	2.719	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041125\
Data File : VX045749.D
Acq On : 11 Apr 2025 18:10
Operator : JC/MD
Sample : Q1790-04 20X
Misc : 6.84g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-871-G1-SO-23.0-23.5-041025

Quant Time: Apr 12 00:21:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

