

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
 Data File : VX008882.D
 Acq On : 12 Apr 2019 13:34
 Operator : JC/SP
 Sample : K2366-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20190409

Quant Time: Apr 13 05:26:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	198329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	316164	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	278345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118722	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	131277	45.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.76%	
35) Dibromofluoromethane	5.49	113	93611	46.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.62%	
50) Toluene-d8	8.71	98	397906	49.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.13	95	133247	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
Data File : VX008882.D
Acq On : 12 Apr 2019 13:34
Operator : JC/SP
Sample : K2366-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20190409

Quant Time: Apr 13 05:26:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

