

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
 Data File : VX008905.D
 Acq On : 12 Apr 2019 23:18
 Operator : JC/SP
 Sample : K2311-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0236-COMP-CS-1-ELTRT-DISS

Quant Time: Apr 13 06:25:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	187609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	301073	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	260543	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110865	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	124721	45.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.16%	
35) Dibromofluoromethane	5.50	113	90115	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
50) Toluene-d8	8.71	98	377870	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	11.13	95	124684	45.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.74%	
Target Compounds						
11) Tert butyl alcohol	3.05	59	4027	6.286	ug/l #	93
16) Acetone	2.45	43	31259	19.833	ug/l	99
20) Methylene Chloride	2.86	84	11652	4.417	ug/l	98
70) Styrene	10.71	104	2483	0.384	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
Data File : VX008905.D
Acq On : 12 Apr 2019 23:18
Operator : JC/SP
Sample : K2311-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0236-COMP-CS-1-ELTRT-DISS

Quant Time: Apr 13 06:25:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

