

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\WX041219\
 Data File : VX008916.D
 Acq On : 13 Apr 2019 03:35
 Operator : JC/SP
 Sample : K2381-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-COMP-3

Manual Integrations
 APPROVED

MMDadoda
 4/13/2019 2:12:32 PM

Quant Time: Apr 13 10:30:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	199435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	318754	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	272380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113849	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	129647	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
35) Dibromofluoromethane	5.49	113	94296	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	8.71	98	398843	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
62) 4-Bromofluorobenzene	11.13	95	129478	44.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.98%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.05	59	5802	8.520	ug/l #	84
16) Acetone	2.45	43	1395399	832.862	ug/l	99
17) Carbon Disulfide	2.57	76	33380	4.338	ug/l	96
18) Methyl Acetate	2.78	43	468363	115.264	ug/l	98
20) Methylene Chloride	2.85	84	1453	0.518	ug/l	93
25) 2-Butanone	4.67	43	174151	65.766	ug/l	98
30) Chloroform	5.21	83	2460	0.547	ug/l	98
37) Ethyl Acetate	4.83	43	966962	207.913	ug/l	99
52) Toluene	8.79	92	11838	1.934	ug/l	92
64) Tetrachloroethene	9.34	164	45302	22.448	ug/l	99
67) Ethyl Benzene	10.25	91	5052	0.449	ug/l	96
68) m/p-Xylenes	10.35	106	8099	2.001	ug/l	98
69) o-Xylene	10.69	106	4897	1.248	ug/l	95
73) Isopropylbenzene	11.02	105	4056	0.430	ug/l	98
78) n-propylbenzene	11.36	91	21813	1.986	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	48503	6.124	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	190006	23.693	ug/l	99
85) sec-Butylbenzene	11.94	105	16900	1.867	ug/l	96
86) p-Isopropyltoluene	12.06	119	17057m	2.100	ug/l	
89) n-Butylbenzene	12.38	91	50363	6.611	ug/l #	45
95) Naphthalene	13.83	128	279233	29.805	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
Data File : VX008916.D
Acq On : 13 Apr 2019 03:35
Operator : JC/SP
Sample : K2381-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-COMP-3

Manual Integrations
APPROVED

MMDadoda
4/13/2019 2:12:32 PM

Quant Time: Apr 13 10:30:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

