

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
 Data File : VX008919.D
 Acq On : 13 Apr 2019 04:45
 Operator : JC/SP
 Sample : K2364-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-COMP-2

Quant Time: Apr 13 07:09:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	200650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	318318	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	280986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118920	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	128027	43.75	ug/l	0.00
Spiked Amount			Recovery	=	87.50%	
35) Dibromofluoromethane	5.50	113	93403	45.89	ug/l	0.00
Spiked Amount			Recovery	=	91.78%	
50) Toluene-d8	8.71	98	400616	49.44	ug/l	0.00
Spiked Amount			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.13	95	132231	46.01	ug/l	0.00
Spiked Amount			Recovery	=	92.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.05	59	16419	23.964	ug/l #	93
16) Acetone	2.45	43	882635	523.622	ug/l	100
18) Methyl Acetate	2.78	43	206912	50.613	ug/l	99
20) Methylene Chloride	2.86	84	2962	1.050	ug/l	84
25) 2-Butanone	4.68	43	227509	85.395	ug/l	100
37) Ethyl Acetate	4.84	43	99472	21.417	ug/l	98
40) Benzene	6.15	78	8707	0.840	ug/l	94
52) Toluene	8.79	92	11359	1.859	ug/l	98
68) m/p-Xylenes	10.36	106	3527	0.845	ug/l	89
69) o-Xylene	10.70	106	1542	0.381	ug/l	93
80) 1,3,5-Trimethylbenzene	11.51	105	2748	0.332	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	11801	1.409	ug/l	86
95) Naphthalene	13.83	128	42922	4.386	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX041219\
Data File : VX008919.D
Acq On : 13 Apr 2019 04:45
Operator : JC/SP
Sample : K2364-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-COMP-2

Quant Time: Apr 13 07:09:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

