

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
 Data File : VX028056.D
 Acq On : 12 Apr 2022 15:05
 Operator : JC/MD
 Sample : VX0412WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0412WBL01

Quant Time: Apr 13 01:11:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	243089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	413296	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	385606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	172707	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158124	47.946	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.900%
35) Dibromofluoromethane	5.385	113	134240	49.568	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.140%
50) Toluene-d8	8.646	98	501776	49.679	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	165330	43.839	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
Data File : VX028056.D
Acq On : 12 Apr 2022 15:05
Operator : JC/MD
Sample : VX0412WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0412WBL01

Quant Time: Apr 13 01:11:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

