

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
 Data File : VX028059.D
 Acq On : 12 Apr 2022 16:24
 Operator : JC/MD
 Sample : N2334-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWO-INCH-WELLDL

Quant Time: Apr 13 01:12:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	390887	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148542	48.756	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.520%
35) Dibromofluoromethane	5.391	113	128221	50.060	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.120%
50) Toluene-d8	8.653	98	479585	50.204	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.400%
62) 4-Bromofluorobenzene	11.079	95	168257	47.173	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.340%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	5513	4.343	ug/l	97
22) Diisopropyl ether	3.763	45	12469	1.481	ug/l	90
25) 2-Butanone	4.568	43	4272	1.947	ug/l #	85
31) Cyclohexane	5.470	56	108973	26.013	ug/l	98
39) Methylcyclohexane	7.385	83	46061	9.968	ug/l	89
40) Benzene	6.043	78	73686	6.482	ug/l	98
52) Toluene	8.720	92	8909	1.175	ug/l	98
67) Ethyl Benzene	10.195	91	31456	2.187	ug/l	99
68) m/p-Xylenes	10.305	106	11284	1.878	ug/l	97
73) Isopropylbenzene	10.963	105	235262	18.267	ug/l	99
78) n-propylbenzene	11.305	91	673048	44.965	ug/l	99
85) sec-Butylbenzene	11.890	105	23752	1.729	ug/l	91
89) n-Butylbenzene	12.335	91	30774	3.088	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
Data File : VX028059.D
Acq On : 12 Apr 2022 16:24
Operator : JC/MD
Sample : N2334-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWO-INCH-WELLDL

Quant Time: Apr 13 01:12:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

