

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
 Data File : VX028060.D
 Acq On : 12 Apr 2022 16:47
 Operator : JC/MD
 Sample : N2334-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB040622

Quant Time: Apr 13 01:13:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	223147	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	383749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	145837	48.173	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.340%
35) Dibromofluoromethane	5.385	113	124571	49.539	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.080%
50) Toluene-d8	8.653	98	466016	49.691	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.085	95	154301	44.065	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.140%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2944	2.334	ug/l	94
20) Methylene Chloride	2.794	84	1344	0.535	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
Data File : VX028060.D
Acq On : 12 Apr 2022 16:47
Operator : JC/MD
Sample : N2334-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB040622

Quant Time: Apr 13 01:13:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

