

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041222\  
 Data File : VX028062.D  
 Acq On : 12 Apr 2022 17:33  
 Operator : JC/MD  
 Sample : N2334-14  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FB040522

Quant Time: Apr 13 01:13:30 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 12 22:07:23 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.556	168	215820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377108	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	355808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142425	48.643	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.280%
35) Dibromofluoromethane	5.385	113	122209	49.456	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.920%
50) Toluene-d8	8.653	98	459504	49.859	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.720%
62) 4-Bromofluorobenzene	11.085	95	152369	44.280	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.560%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2188	1.794	ug/l #	74
20) Methylene Chloride	2.788	84	980	0.403	ug/l #	76
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041222\  
 Data File : VX028062.D  
 Acq On : 12 Apr 2022 17:33  
 Operator : JC/MD  
 Sample : N2334-14  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FB040522

Quant Time: Apr 13 01:13:30 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 12 22:07:23 2022  
 Response via : Initial Calibration

