

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
 Data File : VX028069.D
 Acq On : 12 Apr 2022 20:16
 Operator : JC/MD
 Sample : N2333-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14R

Quant Time: Apr 13 01:14:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	229247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	398086	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	367042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152155	48.922	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.840%
35) Dibromofluoromethane	5.391	113	131924	50.574	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.140%
50) Toluene-d8	8.653	98	483264	49.674	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	158924	43.751	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.500%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3615	2.790	ug/l #	85
84) 1,2,4-Trimethylbenzene	11.756	105	2667	0.265	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
Data File : VX028069.D
Acq On : 12 Apr 2022 20:16
Operator : JC/MD
Sample : N2333-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14R

Quant Time: Apr 13 01:14:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

