

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041222\  
 Data File : VX028053.D  
 Acq On : 12 Apr 2022 13:29  
 Operator : JC/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX041222

Quant Time: Apr 13 01:10:20 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 12 22:07:23 2022  
 Response via : Initial Calibration

### Manual Integrations APPROVED

Reviewed By : John Carlone 04/13/2022  
 Supervised By : Mahesh Dadoda 04/13/2022

| Compound                     | R.T.           | QIon | Response            | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|---------------------|---------|--------|----------|
| -----                        |                |      |                     |         |        |          |
| Internal Standards           |                |      |                     |         |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 204982              | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 337461              | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 322100              | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 182270              | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                     |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 139905              | 50.308  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = 100.620% |         |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 113036              | 51.118  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = 102.240% |         |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 440179              | 53.374  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = 106.740% |         |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 167184              | 54.293  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = 108.580% |         |        |          |
| Target Compounds             |                |      |                     |         |        |          |
|                              |                |      |                     |         |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 92321               | 46.386  | ug/l   | 99       |
| 3) Chloromethane             | 1.288          | 50   | 101718              | 46.179  | ug/l   | 100      |
| 4) Vinyl Chloride            | 1.374          | 62   | 125939              | 47.557  | ug/l   | 99       |
| 5) Bromomethane              | 1.599          | 94   | 156156              | 55.978  | ug/l   | 99       |
| 6) Chloroethane              | 1.678          | 64   | 97956               | 45.239  | ug/l   | 98       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 246262              | 47.418  | ug/l   | 96       |
| 8) Diethyl Ether             | 2.136          | 74   | 90661               | 45.552  | ug/l   | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 93234               | 45.125  | ug/l   | 97       |
| 10) Methyl Iodide            | 2.453          | 142  | 112613              | 46.804  | ug/l   | 96       |
| 11) Tert butyl alcohol       | 2.965          | 59   | 147412              | 228.360 | ug/l   | 100      |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 91253               | 45.102  | ug/l   | 97       |
| 13) Acrolein                 | 2.233          | 56   | 82315               | 240.539 | ug/l   | 100      |
| 14) Allyl chloride           | 2.666          | 41   | 162434              | 45.362  | ug/l   | 88       |
| 15) Acrylonitrile            | 3.062          | 53   | 319739              | 239.033 | ug/l   | 99       |
| 16) Acetone                  | 2.379          | 43   | 266770              | 230.241 | ug/l   | 90       |
| 17) Carbon Disulfide         | 2.507          | 76   | 258817              | 48.677  | ug/l   | 100      |
| 18) Methyl Acetate           | 2.703          | 43   | 140489              | 45.440  | ug/l   | 94       |
| 19) Methyl tert-butyl Ether  | 3.117          | 73   | 368206              | 48.457  | ug/l   | 99       |
| 20) Methylene Chloride       | 2.788          | 84   | 107369              | 46.525  | ug/l   | 89       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 112083              | 46.245  | ug/l   | 98       |
| 22) Diisopropyl ether        | 3.763          | 45   | 371738              | 48.377  | ug/l # | 88       |
| 23) Vinyl Acetate            | 3.721          | 43   | 1707420             | 250.067 | ug/l   | 94       |
| 24) 1,1-Dichloroethane       | 3.611          | 63   | 198713              | 47.771  | ug/l   | 99       |
| 25) 2-Butanone               | 4.556          | 43   | 475195              | 237.268 | ug/l   | 92       |
| 26) 2,2-Dichloropropane      | 4.477          | 77   | 187195              | 48.473  | ug/l   | 98       |
| 27) cis-1,2-Dichloroethene   | 4.495          | 96   | 128800              | 47.144  | ug/l   | 99       |
| 28) Bromochloromethane       | 4.903          | 49   | 82248               | 46.000  | ug/l   | 90       |
| 29) Tetrahydrofuran          | 5.007          | 42   | 308568              | 235.723 | ug/l   | 91       |
| 30) Chloroform               | 5.098          | 83   | 219084              | 47.363  | ug/l   | 98       |
| 31) Cyclohexane              | 5.476          | 56   | 180948              | 47.321  | ug/l   | 92       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 200681              | 47.901  | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 159323              | 47.209  | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.714          | 43   | 187020              | 47.570  | ug/l   | 95       |
| 38) Carbon Tetrachloride     | 5.678          | 117  | 173280              | 48.094  | ug/l   | 99       |
| 39) Methylcyclohexane        | 7.385          | 83   | 192587              | 48.274  | ug/l   | 95       |
| 40) Benzene                  | 6.043          | 78   | 474836              | 48.385  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041222\  
 Data File : VX028053.D  
 Acq On : 12 Apr 2022 13:29  
 Operator : JC/MD  
 Sample : VSTDICV050  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX041222

Manual Integrations  
 APPROVED

Reviewed By : John Carlone 04/13/2022  
 Supervised By : Mahesh Dadoda 04/13/2022

Quant Time: Apr 13 01:10:20 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 12 22:07:23 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.928  | 41   | 97836    | 46.775  | ug/l   | 91       |
| 42) 1,2-Dichloroethane        | 6.092  | 62   | 179490   | 48.436  | ug/l   | 97       |
| 43) Isopropyl Acetate         | 6.342  | 43   | 312887   | 47.770  | ug/l   | 92       |
| 44) Trichloroethene           | 7.129  | 130  | 128141   | 49.103  | ug/l   | 100      |
| 45) 1,2-Dichloropropane       | 7.433  | 63   | 116876   | 47.258  | ug/l   | 99       |
| 46) Dibromomethane            | 7.586  | 93   | 87545    | 47.982  | ug/l   | 94       |
| 47) Bromodichloromethane      | 7.824  | 83   | 169697   | 48.124  | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.696  | 41   | 137477   | 45.978  | ug/l # | 82       |
| 49) 1,4-Dioxane               | 7.659  | 88   | 63396    | 949.743 | ug/l   | 95       |
| 51) 4-Methyl-2-Pentanone      | 8.573  | 43   | 989517   | 236.947 | ug/l   | 91       |
| 52) Toluene                   | 8.720  | 92   | 316544   | 48.347  | ug/l   | 100      |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 201508   | 50.593  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 8.366  | 75   | 202462   | 49.009  | ug/l # | 83       |
| 55) 1,1,2-Trichloroethane     | 9.153  | 97   | 126311   | 47.590  | ug/l   | 97       |
| 56) Ethyl methacrylate        | 9.116  | 69   | 204290   | 48.967  | ug/l # | 80       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 208190   | 47.316  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.244  | 63   | 475589   | 252.564 | ug/l   | 99       |
| 59) 2-Hexanone                | 9.433  | 43   | 793844   | 242.718 | ug/l   | 87       |
| 60) Dibromochloromethane      | 9.525  | 129  | 140104   | 49.380  | ug/l   | 96       |
| 61) 1,2-Dibromoethane         | 9.610  | 107  | 140123   | 48.474  | ug/l   | 100      |
| 64) Tetrachloroethene         | 9.275  | 164  | 120129   | 47.006  | ug/l   | 97       |
| 65) Chlorobenzene             | 10.079 | 112  | 340357   | 48.428  | ug/l   | 100      |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 125133   | 48.188  | ug/l   | 99       |
| 67) Ethyl Benzene             | 10.195 | 91   | 619228   | 50.034  | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 489037   | 94.629  | ug/l   | 99       |
| 69) o-Xylene                  | 10.646 | 106  | 241301   | 48.033  | ug/l   | 99       |
| 70) Styrene                   | 10.658 | 104  | 409113   | 42.107  | ug/l   | 98       |
| 71) Bromoform                 | 10.805 | 173  | 106407   | 43.082  | ug/l   | 99       |
| 73) Isopropylbenzene          | 10.963 | 105  | 634906   | 49.033  | ug/l   | 100      |
| 74) N-amyl acetate            | 10.841 | 43   | 292673   | 52.300  | ug/l # | 88       |
| 75) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 218393   | 48.279  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane    | 11.238 | 75   | 199319m  | 49.120  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 152224   | 48.124  | ug/l   | 99       |
| 78) n-propylbenzene           | 11.305 | 91   | 751780   | 49.957  | ug/l   | 100      |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 438095   | 48.005  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 560035   | 50.069  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.018 | 75   | 73982    | 51.351  | ug/l   | 91       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 532319   | 49.550  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.719 | 119  | 543047   | 50.180  | ug/l   | 95       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 557063   | 50.870  | ug/l   | 100      |
| 85) sec-Butylbenzene          | 11.890 | 105  | 684335   | 49.546  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 587028   | 50.157  | ug/l   | 97       |
| 87) 1,3-Dichlorobenzene       | 11.969 | 146  | 304684   | 49.434  | ug/l   | 100      |
| 88) 1,4-Dichlorobenzene       | 12.042 | 146  | 314752   | 48.682  | ug/l   | 99       |
| 89) n-Butylbenzene            | 12.335 | 91   | 521205   | 52.029  | ug/l   | 97       |
| 90) Hexachloroethane          | 12.542 | 117  | 100197   | 50.953  | ug/l   | 92       |
| 91) 1,2-Dichlorobenzene       | 12.335 | 146  | 303915   | 49.348  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 50994    | 51.241  | ug/l   | 92       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 177166   | 51.445  | ug/l   | 99       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 70125    | 50.464  | ug/l   | 96       |
| 95) Naphthalene               | 13.780 | 128  | 630418   | 52.414  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 175991   | 51.639  | ug/l   | 96       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041222\  
Data File : VX028053.D  
Acq On : 12 Apr 2022 13:29  
Operator : JC/MD  
Sample : VSTDICV050  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222

Manual Integrations  
APPROVED

Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022

Quant Time: Apr 13 01:10:20 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 12 22:07:23 2022  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

-----  
(#) = qualifier out of range (m) = manual integration (+) = signals summed

## Manual Integrations APPROVED


Quant Time: Apr 13 01:10:20 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 12 22:07:23 2022  
Response via : Initial Calibration





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 71  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion:168 Resp: 20498  
Ion Ratio Lower Upper  
168 100  
99 54.1 41.7 62.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#2  
Dichlorodifluoromethane  
Concen: 46.386 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion: 85 Resp: 92321  
Ion Ratio Lower Upper  
85 100  
87 32.7 16.2 48.6





#3  
Chloromethane  
Concen: 46.179 ug/l  
RT: 1.288 min Scan# 33  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 50 Resp: 101718  
Ion Ratio Lower Upper  
50 100  
52 32.8 26.4 39.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#4  
Vinyl Chloride  
Concen: 47.557 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 62 Resp: 125939  
Ion Ratio Lower Upper  
62 100  
64 32.2 25.5 38.3





#5  
Bromomethane  
Concen: 55.978 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVX041222




Tgt Ion: 94 Resp: 156150  
Ion Ratio Lower Upper  
94 100  
96 93.2 75.3 112.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#6  
Chloroethane  
Concen: 45.239 ug/l  
RT: 1.678 min Scan# 97  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 64 Resp: 97956  
Ion Ratio Lower Upper  
64 100  
66 32.6 25.3 37.9





#7  
Trichlorofluoromethane  
Concen: 47.418 ug/l  
RT: 1.886 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion:101 Resp: 246262  
Ion Ratio Lower Upper  
101 100  
103 66.2 50.5 75.7




Manual Integrations  
APPROVED

Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#8  
Diethyl Ether  
Concen: 45.552 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Tgt Ion: 74 Resp: 90661  
Ion Ratio Lower Upper  
74 100  
45 86.0 42.0 126.0











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 45.125 ug/l  
RT: 2.325 min Scan# 203  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222

Tgt Ion:101 Resp: 93234  
Ion Ratio Lower Upper  
101 100  
85 46.4 33.5 50.3  
151 76.2 61.8 92.6




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#10  
Methyl Iodide  
Concen: 46.804 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Tgt Ion:142 Resp: 112613  
Ion Ratio Lower Upper  
142 100  
127 45.5 33.9 50.9  
141 14.2 11.4 17.2





#11

Tert butyl alcohol

Concen: 228.360 ug/l

RT: 2.965 min Scan# 308

Delta R.T. -0.000 min

Lab File: VX028053.D

Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222

Tgt Ion: 59 Resp: 14741

Ion Ratio Lower Upper

59 100





57 10.2 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/13/2022

Supervised By :Mahesh Dadoda 04/13/2022



#12

1,1-Dichloroethene

Concen: 45.102 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX028053.D

Acq: 12 Apr 2022 13:29




Tgt Ion: 96 Resp: 91253


Ion Ratio Lower Upper

96 100

61 161.1 124.6 187.0


98 64.5 51.0 76.4





#13  
Acrolein  
Concen: 240.539 ug/l  
RT: 2.233 min Scan# 188  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 56 Resp: 82315  
Ion Ratio Lower Upper  
56 100  
55 68.9 55.0 82.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#14  
Allyl chloride  
Concen: 45.362 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 41 Resp: 162434  
Ion Ratio Lower Upper  
41 100  
39 74.3 50.4 75.6  
76 36.3 32.7 49.1





#15  
Acrylonitrile  
Concen: 239.033 ug/l  
RT: 3.062 min Scan# 31  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 53 Resp: 319739  
Ion Ratio Lower Upper  
53 100  
52 83.0 65.7 98.5  
51 36.4 28.4 42.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#16  
Acetone  
Concen: 230.241 ug/l  
RT: 2.379 min Scan# 212  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion: 43 Resp: 266770  
Ion Ratio Lower Upper  
43 100  
58 29.9 28.8 43.2






#17  
Carbon Disulfide  
Concen: 48.677 ug/l  
RT: 2.507 min Scan# 211  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222






Tgt Ion: 76 Resp: 25881  
Ion Ratio Lower Upper  
76 100  
78 9.1 7.2 10.8


Manual Integrations

APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#18  
Methyl Acetate  
Concen: 45.440 ug/l  
RT: 2.703 min Scan# 265  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 43 Resp: 140489  
Ion Ratio Lower Upper  
43 100  
74 25.1 22.8 34.2





#19  
Methyl tert-butyl Ether  
Concen: 48.457 ug/l  
RT: 3.117 min Scan# 311  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion: 73 Resp: 368200  
Ion Ratio Lower Upper  
73 100  
57 20.8 16.8 25.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#20  
Methylene Chloride  
Concen: 46.525 ug/l  
RT: 2.788 min Scan# 279  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion: 84 Resp: 107369  
Ion Ratio Lower Upper  
84 100  
49 126.2 85.5 128.3  
51 37.9 26.6 39.8  
86 63.4 51.5 77.3





#21  
trans-1,2-Dichloroethene  
Concen: 46.245 ug/l  
RT: 3.093 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 96 Resp: 11208  
Ion Ratio Lower Upper  
96 100  
61 131.2 105.4 158.2  
98 65.5 50.0 75.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#22  
Diisopropyl ether  
Concen: 48.377 ug/l  
RT: 3.763 min Scan# 439  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 45 Resp: 371738  
Ion Ratio Lower Upper  
45 100  
43 76.5 50.8 76.2#  
87 30.1 26.6 39.8  
59 12.1 10.5 15.7





#23  
 Vinyl Acetate  
 Concen: 250.067 ug/l  
 RT: 3.721 min Scan# 41  
 Delta R.T. -0.000 min  
 Lab File: VX028053.D  
 Acq: 12 Apr 2022 13:29




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX041222




Tgt Ion: 43 Resp: 1707420  
 Ion Ratio Lower Upper  
 43 100  
 86 11.1 10.8 16.2

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 04/13/2022  
 Supervised By :Mahesh Dadoda 04/13/2022




#24  
 1,1-Dichloroethane  
 Concen: 47.771 ug/l  
 RT: 3.611 min Scan# 414  
 Delta R.T. -0.000 min  
 Lab File: VX028053.D  
 Acq: 12 Apr 2022 13:29



Tgt Ion: 63 Resp: 198713  
 Ion Ratio Lower Upper  
 63 100  
 98 7.5 3.9 11.7  
 100 5.3 2.4 7.2








#25  
2-Butanone  
Concen: 237.268 ug/l  
RT: 4.556 min Scan# 50  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 43 Resp: 475195  
Ion Ratio Lower Upper  
43 100  
72 26.2 24.4 36.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#26  
2,2-Dichloropropane  
Concen: 48.473 ug/l  
RT: 4.477 min Scan# 556  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 77 Resp: 187195  
Ion Ratio Lower Upper  
77 100  
97 22.5 11.8 35.4





#27  
 cis-1,2-Dichloroethene  
 Concen: 47.144 ug/l  
 RT: 4.495 min Scan# 558  
 Delta R.T. 0.006 min  
 Lab File: VX028053.D  
 Acq: 12 Apr 2022 13:29




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX041222




Tgt Ion: 96 Resp: 128800  
 Ion Ratio Lower Upper  
 96 100  
 61 143.3 0.0 283.8  
 98 64.0 0.0 129.8

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 04/13/2022  
 Supervised By :Mahesh Dadoda 04/13/2022




#28  
 Bromochloromethane  
 Concen: 46.000 ug/l  
 RT: 4.903 min Scan# 626  
 Delta R.T. -0.001 min  
 Lab File: VX028053.D  
 Acq: 12 Apr 2022 13:29



Tgt Ion: 49 Resp: 82248  
 Ion Ratio Lower Upper  
 49 100  
 129 2.2 0.0 4.4  
 130 79.0 70.7 106.1





#29

Tetrahydrofuran

Concen: 235.723 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX028053.D


Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222



Tgt Ion: 42 Resp: 30856

Ion Ratio Lower Upper

42 100

72 47.9 44.1 66.1




71 44.4 40.5 60.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/13/2022

Supervised By :Mahesh Dadoda 04/13/2022



#30

Chloroform

Concen: 47.363 ug/l

RT: 5.098 min Scan# 658

Delta R.T. -0.001 min

Lab File: VX028053.D




Acq: 12 Apr 2022 13:29


Tgt Ion: 83 Resp: 219084

Ion Ratio Lower Upper

83 100


85 64.0 52.7 79.1





#31  
Cyclohexane  
Concen: 47.321 ug/l  
RT: 5.476 min Scan# 720  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion: 56 Resp: 180948  
Ion Ratio Lower Upper  
56 100  
69 30.4 26.7 40.1  
84 88.1 77.2 115.8




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#32  
1,1,1-Trichloroethane  
Concen: 47.901 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion: 97 Resp: 200681  
Ion Ratio Lower Upper  
97 100  
99 63.6 51.8 77.6  
61 43.4 34.0 51.0





#33  
1,2-Dichloroethane-d4  
Concen: 50.308 ug/l  
RT: 5.958 min Scan# 799  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 65 Resp: 139908  
Ion Ratio Lower Upper  
65 100  
67 51.8 0.0 107.4

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.763 min Scan# 931  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:114 Resp: 337461  
Ion Ratio Lower Upper  
114 100  
63 19.6 0.0 36.8  
88 15.1 0.0 32.4





#35

Dibromofluoromethane

Concen: 51.118 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX028053.D

Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222

Tgt Ion: 113 Resp: 113030





| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 113 | 100   |       |       |
| 111 | 103.3 | 82.8  | 124.2 |
| 192 | 20.3  | 16.9  | 25.3  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/13/2022

Supervised By :Mahesh Dadoda 04/13/2022



#36

1,1-Dichloropropene

Concen: 47.209 ug/l

RT: 5.696 min Scan# 756




Delta R.T. -0.000 min


Lab File: VX028053.D

Acq: 12 Apr 2022 13:29

Tgt Ion: 75 Resp: 159323


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 75  | 100   |       |       |
| 110 | 35.4  | 18.3  | 54.8  |
| 77  | 30.4  | 24.8  | 37.2  |





#37  
Ethyl Acetate  
Concen: 47.570 ug/l  
RT: 4.714 min Scan# 595  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 43 Resp: 187020  
Ion Ratio Lower Upper  
43 100  
61 14.0 12.6 19.0  
70 10.7 10.2 15.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#38  
Carbon Tetrachloride  
Concen: 48.094 ug/l  
RT: 5.678 min Scan# 753  
Delta R.T. -0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:117 Resp: 173280  
Ion Ratio Lower Upper  
117 100  
119 96.2 77.3 115.9  
121 29.5 24.2 36.2





#39

Methylcyclohexane

Concen: 48.274 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX028053.D


Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222





|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 83    | Resp: | 19258 |
| Ion      | Ratio | Lower | Upper |
| 83       | 100   |       |       |
| 55       | 72.8  | 54.6  | 81.8  |
| 98       | 50.0  | 37.4  | 56.2  |

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/13/2022

Supervised By :Mahesh Dadoda 04/13/2022



#40

Benzene

Concen: 48.385 ug/l




RT: 6.043 min Scan# 813

Delta R.T. 0.006 min


Lab File: VX028053.D

Acq: 12 Apr 2022 13:29

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 78    | Resp: | 474836 |
| Ion      | Ratio | Lower | Upper  |
| 78       | 100   |       |        |
| 77       | 23.6  | 19.3  | 28.9   |








#41  
Methacrylonitrile  
Concen: 46.775 ug/l  
RT: 4.928 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 41 Resp: 97830  
Ion Ratio Lower Upper  
41 100  
39 58.0 43.0 64.4  
67 74.8 70.2 105.4  
52 32.5 26.0 39.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#42  
1,2-Dichloroethane  
Concen: 48.436 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 62 Resp: 179490  
Ion Ratio Lower Upper  
62 100  
98 9.0 0.0 20.4





#43  
Isopropyl Acetate  
Concen: 47.770 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 43 Resp: 31288  
Ion Ratio Lower Upper  
43 100  
61 18.9 18.5 27.7  
87 13.6 13.0 19.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#44  
Trichloroethene  
Concen: 49.103 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:130 Resp: 128141  
Ion Ratio Lower Upper  
130 100  
95 94.5 0.0 188.8





#45

1,2-Dichloropropane

Concen: 47.258 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.001 min

Lab File: VX028053.D

Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222

Tgt Ion: 63 Resp: 116870

Ion Ratio Lower Upper

63 100





65 31.4 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/13/2022

Supervised By :Mahesh Dadoda 04/13/2022



#46

Dibromomethane

Concen: 47.982 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX028053.D

Acq: 12 Apr 2022 13:29




Tgt Ion: 93 Resp: 87545


Ion Ratio Lower Upper

93 100

95 81.6 65.8 98.6


174 97.0 85.4 128.2





#47  
Bromodichloromethane  
Concen: 48.124 ug/l  
RT: 7.824 min Scan# 1105  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 83 Resp: 169697  
Ion Ratio Lower Upper  
83 100  
85 64.1 52.2 78.2  
127 8.5 6.8 10.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#48  
Methyl methacrylate  
Concen: 45.978 ug/l  
RT: 7.696 min Scan# 1084  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 41 Resp: 137477  
Ion Ratio Lower Upper  
41 100  
69 86.1 82.5 123.7  
39 68.8 43.8 65.8





#49  
1,4-Dioxane  
Concen: 949.743 ug/l  
RT: 7.659 min Scan# 1078  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion: 88 Resp: 63390  
Ion Ratio Lower Upper  
88 100  
43 31.4 21.9 32.9  
58 63.6 53.7 80.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#50  
Toluene-d8  
Concen: 53.374 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Tgt Ion: 98 Resp: 440179  
Ion Ratio Lower Upper  
98 100  
100 64.9 52.6 78.8









#51  
4-Methyl-2-Pentanone  
Concen: 236.947 ug/l  
RT: 8.573 min Scan# 1228  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Instrument : MSVOA\_X  
ClientSampleId : ICVVX041222

Tgt Ion: 43 Resp: 98951  
Ion Ratio Lower Upper  
43 100  
58 39.0 36.1 54.1




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#52  
Toluene  
Concen: 48.347 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion: 92 Resp: 316544  
Ion Ratio Lower Upper  
92 100  
91 171.9 137.5 206.3





#53  
t-1,3-Dichloropropene  
Concen: 50.593 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 75 Resp: 201508  
Ion Ratio Lower Upper  
75 100  
77 30.7 25.4 38.0

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#54  
cis-1,3-Dichloropropene  
Concen: 49.009 ug/l  
RT: 8.366 min Scan# 1194  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 75 Resp: 202462  
Ion Ratio Lower Upper  
75 100  
77 32.7 24.8 37.2  
39 54.1 29.6 44.4#





#55  
1,1,2-Trichloroethane  
Concen: 47.590 ug/l  
RT: 9.153 min Scan# 1323  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29

Instrument : MSVOA\_X  
ClientSampleId : ICVVX041222






Tgt Ion: 97 Resp: 12631


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 86.5  | 66.2  | 99.4  |
| 85  | 55.8  | 42.9  | 64.3  |
| 99  | 62.5  | 49.0  | 73.6  |

Manual Integrations  
APPROVED

Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022





#56  
Ethyl methacrylate  
Concen: 48.967 ug/l  
RT: 9.116 min Scan# 1317  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 69 Resp: 204290

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 65.5  | 43.8  | 65.6  |
| 39  | 43.9  | 22.2  | 33.2  |








#57  
1,3-Dichloropropane  
Concen: 47.316 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 76 Resp: 208190  
Ion Ratio Lower Upper  
76 100  
78 32.5 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#58  
2-Chloroethyl Vinyl ether  
Concen: 252.564 ug/l  
RT: 8.244 min Scan# 1174  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 63 Resp: 475589  
Ion Ratio Lower Upper  
63 100  
106 29.5 23.9 35.9





#59  
2-Hexanone  
Concen: 242.718 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion: 43 Resp: 793844  
Ion Ratio Lower Upper  
43 100  
58 54.0 31.9 95.7




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#60  
Dibromochloromethane  
Concen: 49.380 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion:129 Resp: 140104  
Ion Ratio Lower Upper  
129 100  
127 76.7 40.3 120.8





#61  
1,2-Dibromoethane  
Concen: 48.474 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:107 Resp: 14012  
Ion Ratio Lower Upper  
107 100  
109 92.6 74.4 111.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#62  
4-Bromofluorobenzene  
Concen: 54.293 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 95 Resp: 167184  
Ion Ratio Lower Upper  
95 100  
174 80.8 0.0 163.0  
176 79.4 0.0 155.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion:117 Resp: 322100  
Ion Ratio Lower Upper  
117 100  
82 54.9 41.8 62.8  
119 31.8 24.8 37.2




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#64  
Tetrachloroethene  
Concen: 47.006 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion:164 Resp: 120129  
Ion Ratio Lower Upper  
164 100  
166 122.3 101.2 151.8  
129 90.2 73.0 109.4  
131 87.6 74.2 111.2





#65  
Chlorobenzene  
Concen: 48.428 ug/l  
RT: 10.079 min Scan# 1475  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:112 Resp: 34035  
Ion Ratio Lower Upper  
112 100  
114 32.2 25.8 38.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 48.188 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:131 Resp: 125133  
Ion Ratio Lower Upper  
131 100  
133 97.1 48.3 144.8  
119 67.0 32.6 98.0





#67

Ethyl Benzene

Concen: 50.034 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX028053.D

Acq: 12 Apr 2022 13:29

Instrument :

MSVOA\_X

ClientSampleId :

ICVVX041222

Tgt Ion: 91 Resp: 619228

Ion Ratio Lower Upper

91 100





106 31.5 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 04/13/2022

Supervised By :Mahesh Dadoda 04/13/2022



#68

m/p-Xylenes

Concen: 94.629 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX028053.D




Acq: 12 Apr 2022 13:29


Tgt Ion:106 Resp: 489037

Ion Ratio Lower Upper

106 100


91 202.4 160.4 240.6





#69  
o-Xylene  
Concen: 48.033 ug/l  
RT: 10.646 min Scan# 1567  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:106 Resp: 24130  
Ion Ratio Lower Upper  
106 100  
91 211.0 106.1 318.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#70  
Styrene  
Concen: 42.107 ug/l  
RT: 10.658 min Scan# 1570  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion:104 Resp: 409113  
Ion Ratio Lower Upper  
104 100  
78 50.6 38.6 57.8  
103 54.6 43.8 65.8





#71  
Bromoform  
Concen: 43.082 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:173 Resp: 10640  
Ion Ratio Lower Upper  
173 100  
175 48.9 24.1 72.2  
254 0.1 0.1 0.1

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:152 Resp: 182270  
Ion Ratio Lower Upper  
152 100  
115 85.4 37.5 112.7  
150 175.9 0.0 338.2








#73  
Isopropylbenzene  
Concen: 49.033 ug/l  
RT: 10.963 min Scan# 1620  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222



Tgt Ion:105 Resp: 634900  
Ion Ratio Lower Upper  
105 100  
120 27.1 13.5 40.5




Manual Integrations  
APPROVED


Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#74  
N-ethyl acetate  
Concen: 52.300 ug/l  
RT: 10.841 min Scan# 1600  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29


Tgt Ion: 43 Resp: 292673  
Ion Ratio Lower Upper  
43 100  
70 42.7 42.5 63.7  
55 22.6 23.1 34.7#  
61 23.8 22.8 34.2





#75  
1,1,2,2-Tetrachloroethane  
Concen: 48.279 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 83 Resp: 21839  
Ion Ratio Lower Upper  
83 100  
131 9.9 5.1 15.3  
85 65.0 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#76  
1,2,3-Trichloropropane  
Concen: 49.120 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. -0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 75 Resp: 199319  
Ion Ratio Lower Upper  
75 100  
77 38.9 19.4 58.2





#77  
Bromobenzene  
Concen: 48.124 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 156 Resp: 152224  
Ion Ratio Lower Upper  
156 100  
77 157.4 78.5 235.3  
158 97.3 50.2 150.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#78  
n-propylbenzene  
Concen: 49.957 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 91 Resp: 751780  
Ion Ratio Lower Upper  
91 100  
120 23.3 11.6 34.8





#79  
2-Chlorotoluene  
Concen: 48.005 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 91 Resp: 438095  
Ion Ratio Lower Upper  
91 100  
126 34.5 17.3 51.9

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#80  
1,3,5-Trimethylbenzene  
Concen: 50.069 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion:105 Resp: 560035  
Ion Ratio Lower Upper  
105 100  
120 48.4 24.3 72.9





#81  
trans-1,4-Dichloro-2-butene  
Concen: 51.351 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 75 Resp: 73982  
Ion Ratio Lower Upper  
75 100  
53 98.3 69.0 103.4  
89 47.1 37.0 55.6

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#82  
4-Chlorotoluene  
Concen: 49.550 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 91 Resp: 532319  
Ion Ratio Lower Upper  
91 100  
126 30.4 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 50.180 ug/l  
RT: 11.719 min Scan# 1174  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:119 Resp: 54304  
Ion Ratio Lower Upper  
119 100  
91 58.3 26.9 80.6  
134 23.6 11.4 34.2

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#84  
1,2,4-Trimethylbenzene  
Concen: 50.870 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion:105 Resp: 557063  
Ion Ratio Lower Upper  
105 100  
120 45.1 22.6 67.8





#85  
 sec-Butylbenzene  
 Concen: 49.546 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. -0.000 min  
 Lab File: VX028053.D  
 Acq: 12 Apr 2022 13:29




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ICVVX041222




Tgt Ion:105 Resp: 684335  
 Ion Ratio Lower Upper  
 105 100  
 134 20.4 10.2 30.4

Manual Integrations  
 APPROVED



Reviewed By :John Carlone 04/13/2022  
 Supervised By :Mahesh Dadoda 04/13/2022




#86  
 p-Isopropyltoluene  
 Concen: 50.157 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. -0.000 min  
 Lab File: VX028053.D  
 Acq: 12 Apr 2022 13:29




Tgt Ion:119 Resp: 587028  
 Ion Ratio Lower Upper  
 119 100  
 134 27.0 13.0 39.0  
 91 24.7 11.3 33.9





#87  
1,3-Dichlorobenzene  
Concen: 49.434 ug/l  
RT: 11.969 min Scan# 1785  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:146 Resp: 304684  
Ion Ratio Lower Upper  
146 100  
111 39.6 19.7 59.1  
148 63.4 31.6 94.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#88  
1,4-Dichlorobenzene  
Concen: 48.682 ug/l  
RT: 12.042 min Scan# 1797  
Delta R.T. -0.001 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:146 Resp: 314752  
Ion Ratio Lower Upper  
146 100  
111 38.8 19.2 57.6  
148 64.2 32.3 96.9








#89  
n-Butylbenzene  
Concen: 52.029 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 91 Resp: 521209  
Ion Ratio Lower Upper  
91 100  
92 53.1 27.8 83.3  
134 26.1 13.3 39.8

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#90  
Hexachloroethane  
Concen: 50.953 ug/l  
RT: 12.542 min Scan# 1879  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion:117 Resp: 100197  
Ion Ratio Lower Upper  
117 100  
201 72.1 39.5 118.5





#91  
1,2-Dichlorobenzene  
Concen: 49.348 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion: 146 Resp: 303915  
Ion Ratio Lower Upper  
146 100  
111 41.7 21.0 63.0  
148 64.5 32.6 97.7

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 51.241 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion: 75 Resp: 50994  
Ion Ratio Lower Upper  
75 100  
155 85.4 47.0 141.2  
157 108.2 57.9 173.6





#93  
1,2,4-Trichlorobenzene  
Concen: 51.445 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:180 Resp: 177160  
Ion Ratio Lower Upper  
180 100  
182 93.9 47.7 143.1  
145 30.3 15.2 45.5

Manual Integrations  
APPROVED



Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022




#94  
Hexachlorobutadiene  
Concen: 50.464 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Tgt Ion:225 Resp: 70125  
Ion Ratio Lower Upper  
225 100  
223 64.1 34.2 102.5  
227 64.9 33.9 101.7





#95  
Naphthalene  
Concen: 52.414 ug/l  
RT: 13.780 min Scan# 2081  
Delta R.T. 0.006 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29




Instrument :  
MSVOA\_X  
ClientSampleId :  
ICVVX041222




Tgt Ion:128 Resp: 630418  
Ion Ratio Lower Upper  
128 100  
127 13.0 10.2 15.4  
129 11.1 8.7 13.1

Manual Integrations  
APPROVED

Reviewed By :John Carlone 04/13/2022  
Supervised By :Mahesh Dadoda 04/13/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 51.639 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. -0.000 min  
Lab File: VX028053.D  
Acq: 12 Apr 2022 13:29



Tgt Ion:180 Resp: 175991  
Ion Ratio Lower Upper  
180 100  
182 94.9 50.0 149.8  
145 31.0 16.2 48.5

