



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041222\
 Data File : VX028045.D
 Acq On : 12 Apr 2022 09:52
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Title : SW846 8260
 Last Update : Tue Apr 12 22:07:23 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	8621	PASS
75	95	30	60	56.5	22545	PASS
95	95	100	100	100.0	39907	PASS
96	95	5	9	6.6	2641	PASS
173	174	0.00	2	1.1	380	PASS
174	95	50	100	88.5	35315	PASS
175	174	5	9	6.9	2448	PASS
176	174	95	101	97.8	34541	PASS
177	176	5	9	6.5	2234	PASS