

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
 Data File : VX035010.D
 Acq On : 11 Apr 2023 15:34
 Operator : JC/MD
 Sample : 02271-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2-040723

Quant Time: Apr 12 05:50:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	286588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	484583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	431331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	200945	45.847	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.700%
35) Dibromofluoromethane	5.385	113	154122	48.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	8.647	98	581432	49.148	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	212308	46.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.760%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.501	96	2191	0.576	ug/l	86
89) n-Butylbenzene	12.335	91	2239	0.239	ug/l	90
93) 1,2,4-Trichlorobenzene	13.591	180	1259	0.365	ug/l	92
94) Hexachlorobutadiene	13.719	225	1586	1.046	ug/l	95
96) 1,2,3-Trichlorobenzene	13.963	180	1270	0.365	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
Data File : VX035010.D
Acq On : 11 Apr 2023 15:34
Operator : JC/MD
Sample : 02271-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2-040723

Quant Time: Apr 12 05:50:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

