

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
 Data File : VX035011.D
 Acq On : 11 Apr 2023 15:57
 Operator : JC/MD
 Sample : 02271-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4-040723

Quant Time: Apr 12 05:51:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	279963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	482833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	430383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	191290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197719	46.178	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.360%
35) Dibromofluoromethane	5.385	113	150275	47.056	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.120%
50) Toluene-d8	8.647	98	573445	48.648	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.079	95	214928	47.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.240%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	17353	4.672	ug/l	88
30) Chloroform	5.105	83	2449	0.394	ug/l	100
44) Trichloroethene	7.135	130	4193	1.209	ug/l	94
64) Tetrachloroethene	9.275	164	6877	2.233	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
Data File : VX035011.D
Acq On : 11 Apr 2023 15:57
Operator : JC/MD
Sample : 02271-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4-040723

Quant Time: Apr 12 05:51:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

