

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
 Data File : VX035018.D
 Acq On : 11 Apr 2023 18:39
 Operator : JC/MD
 Sample : 02263-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Apr 12 05:53:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	265314	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	446653	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	190728	47.005	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.020%
35) Dibromofluoromethane	5.385	113	140998	47.727	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	8.647	98	538959	49.426	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.079	95	196532	46.575	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.140%
Target Compounds						
20) Methylene Chloride	2.794	84	2035	0.626	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
Data File : VX035018.D
Acq On : 11 Apr 2023 18:39
Operator : JC/MD
Sample : 02263-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Apr 12 05:53:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

