

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
 Data File : VX035020.D
 Acq On : 11 Apr 2023 19:26
 Operator : JC/MD
 Sample : 02272-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2-040623

Quant Time: Apr 12 05:54:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	259989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	451007	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	406982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	193503	48.666	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.340%
35) Dibromofluoromethane	5.385	113	142798	47.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.740%
50) Toluene-d8	8.647	98	539299	48.980	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	201390	47.265	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.540%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3645	2.257	ug/l	93
39) Methylcyclohexane	7.385	83	1475	0.314	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
Data File : VX035020.D
Acq On : 11 Apr 2023 19:26
Operator : JC/MD
Sample : 02272-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2-040623

Quant Time: Apr 12 05:54:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

