

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
 Data File : VX035027.D
 Acq On : 11 Apr 2023 22:08
 Operator : JC/MD
 Sample : 02272-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11-040623

Quant Time: Apr 12 05:57:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	267966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	463648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	423055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	194767	47.526	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.060%
35) Dibromofluoromethane	5.385	113	146352	47.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.440%
50) Toluene-d8	8.647	98	558321	49.325	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.660%
62) 4-Bromofluorobenzene	11.079	95	210302	48.011	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
Data File : VX035027.D
Acq On : 11 Apr 2023 22:08
Operator : JC/MD
Sample : 02272-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11-040623

Quant Time: Apr 12 05:57:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

