


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041223\
Data File : VX035004.D
Acq On : 11 Apr 2023 12:30
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Title : SW846 8260
Last Update : Sat Apr 01 01:50:18 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	18699	PASS
75	95	30	60	51.6	48069	PASS
95	95	100	100	100.0	93125	PASS
96	95	5	9	6.8	6315	PASS
173	174	0.00	2	0.8	505	PASS
174	95	50	100	69.3	64581	PASS
175	174	5	9	7.6	4911	PASS
176	174	95	101	96.4	62285	PASS
177	176	5	9	6.9	4291	PASS