

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041019.D
 Acq On : 12 Apr 2024 10:59
 Operator : JC/MD
 Sample : P2048-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 13 01:37:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:36:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	260928	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	250862	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	131414	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	88326	47.592	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.180%
7) Chloroethane-d5	1.672	69	75774	50.062	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.120%
11) 1,1-Dichloroethene-d2	2.306	65	38017	47.980	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.960%
21) 2-Butanone-d5	4.458	46	119747	118.450	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.450%
24) Chloroform-d	5.056	84	169015	48.808	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.620%
26) 1,2-Dichloroethane-d4	5.958	65	111431	49.470	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.940%
32) Benzene-d6	5.976	84	339989	49.921	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.840%
36) 1,2-Dichloropropane-d6	7.305	67	99641	50.165	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.320%
41) Toluene-d8	8.647	98	318935	47.935	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.860%
43) trans-1,3-Dichloroprop...	8.952	79	45716	45.736	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.480%
47) 2-Hexanone-d5	9.384	63	82396	98.390	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.390%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	146877	49.919	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.840%
66) 1,2-Dichlorobenzene-d4	12.317	152	140642	53.351	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041019.D
Acq On : 12 Apr 2024 10:59
Operator : JC/MD
Sample : P2048-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 13 01:37:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:36:25 2024
Response via : Initial Calibration

