

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041035.D
 Acq On : 12 Apr 2024 17:06
 Operator : JC/MD
 Sample : P2049-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 13 01:41:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:36:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	261950	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	249488	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	131639	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	103212	55.396	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 110.800%		
7) Chloroethane-d5	1.673	69	82605	54.362	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 108.720%		
11) 1,1-Dichloroethene-d2	2.307	65	42572	53.519	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 107.040%		
21) 2-Butanone-d5	4.453	46	126897	125.033	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 125.030%		
24) Chloroform-d	5.056	84	178345	51.301	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.600%		
26) 1,2-Dichloroethane-d4	5.958	65	118414	52.365	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.720%		
32) Benzene-d6	5.977	84	364977	53.886	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 107.780%		
36) 1,2-Dichloropropane-d6	7.306	67	105322	53.317	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 106.640%		
41) Toluene-d8	8.647	98	341326	51.583	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 103.160%		
43) trans-1,3-Dichloroprop...	8.952	79	47189	47.470	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 94.940%		
47) 2-Hexanone-d5	9.385	63	87402	104.943	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 104.940%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	148571	50.773	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 101.540%		
66) 1,2-Dichlorobenzene-d4	12.317	152	142676	54.030	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 108.060%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041035.D
Acq On : 12 Apr 2024 17:06
Operator : JC/MD
Sample : P2049-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 13 01:41:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:36:25 2024
Response via : Initial Calibration

