

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041019.D
 Acq On : 12 Apr 2024 10:59
 Operator : JC/MD
 Sample : P2048-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MC0R60

Quant Time: Apr 13 01:37:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:36:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	260928	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	250862	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	131414	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	88326	47.592	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 95.180%			
7) Chloroethane-d5	1.672	69	75774	50.062	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 100.120%			
11) 1,1-Dichloroethene-d2	2.306	65	38017	47.980	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 95.960%			
21) 2-Butanone-d5	4.458	46	119747	118.450	ug/L	0.01
Spiked Amount 100.000	Range 40 - 130		Recovery = 118.450%			
24) Chloroform-d	5.056	84	169015	48.808	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 97.620%			
26) 1,2-Dichloroethane-d4	5.958	65	111431	49.470	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 98.940%			
32) Benzene-d6	5.976	84	339989	49.921	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 99.840%			
36) 1,2-Dichloropropane-d6	7.305	67	99641	50.165	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 100.320%			
41) Toluene-d8	8.647	98	318935	47.935	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 95.860%			
43) trans-1,3-Dichloroprop...	8.952	79	45716	45.736	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 91.480%			
47) 2-Hexanone-d5	9.384	63	82396	98.390	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 98.390%			
56) 1,1,2,2-Tetrachloroeth...	11.189	84	146877	49.919	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 99.840%			
66) 1,2-Dichlorobenzene-d4	12.317	152	140642	53.351	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 106.700%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041019.D
Acq On : 12 Apr 2024 10:59
Operator : JC/MD
Sample : P2048-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC0R60

Quant Time: Apr 13 01:37:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:36:25 2024
Response via : Initial Calibration

