

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041033.D
 Acq On : 12 Apr 2024 16:20
 Operator : JC/MD
 Sample : P2049-05 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MC0RC5

Quant Time: Apr 13 01:40:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:36:25 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	250293	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	237071	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	129485	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	102031	57.313	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 114.620%		
7) Chloroethane-d5	1.666	69	83406	57.445	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 114.900%		
11) 1,1-Dichloroethene-d2	2.306	65	42237	55.570	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 111.140%		
21) 2-Butanone-d5	4.452	46	129762	133.811	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 133.810%#		
24) Chloroform-d	5.056	84	177185	53.341	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.680%		
26) 1,2-Dichloroethane-d4	5.958	65	116870	54.089	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.180%		
32) Benzene-d6	5.970	84	356787	55.435	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 110.880%		
36) 1,2-Dichloropropane-d6	7.305	67	105693	56.307	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 112.620%		
41) Toluene-d8	8.647	98	332097	52.816	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.640%		
43) trans-1,3-Dichloroprop...	8.952	79	48271	51.101	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 102.200%		
47) 2-Hexanone-d5	9.384	63	89726	113.376	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 113.380%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	152491	54.842	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 109.680%		
66) 1,2-Dichlorobenzene-d4	12.317	152	150015	57.755	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 115.500%		
Target Compounds						
						Qvalue
51) Chlorobenzene	10.079	112	146465	30.682	ug/L	98
64) 1,3-Dichlorobenzene	11.969	146	13522	3.404	ug/L	90
65) 1,4-Dichlorobenzene	12.042	146	51845	12.715	ug/L	97
67) 1,2-Dichlorobenzene	12.335	146	489482	127.033	ug/L	96
70) 1,2,4-trichlorobenzene	13.585	180	65107	22.586	ug/L	100
72) 1,2,3-Trichlorobenzene	13.963	180	11355	4.016	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041033.D
Acq On : 12 Apr 2024 16:20
Operator : JC/MD
Sample : P2049-05 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC0RC5

Quant Time: Apr 13 01:40:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:36:25 2024
Response via : Initial Calibration

