

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041036.D
 Acq On : 12 Apr 2024 17:29
 Operator : JC/MD
 Sample : P2049-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CORH7

Quant Time: Apr 13 01:41:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:36:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	254224	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	243773	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	130053	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	97977	54.184	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 108.360%	
7) Chloroethane-d5	1.672	69	80153	54.351	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 108.700%	
11) 1,1-Dichloroethene-d2	2.306	65	39865	51.639	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 103.280%	
21) 2-Butanone-d5	4.452	46	121042	122.889	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 122.890%	
24) Chloroform-d	5.056	84	169640	50.280	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 100.560%	
26) 1,2-Dichloroethane-d4	5.952	65	112339	51.188	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 102.380%	
32) Benzene-d6	5.976	84	346742	52.393	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 104.780%	
36) 1,2-Dichloropropane-d6	7.305	67	99152	51.370	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 102.740%	
41) Toluene-d8	8.647	98	328885	50.868	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 101.740%	
43) trans-1,3-Dichloroprop...	8.952	79	45983	47.341	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 94.680%	
47) 2-Hexanone-d5	9.384	63	85069	104.536	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 104.540%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	146648	51.291	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 102.580%	
66) 1,2-Dichlorobenzene-d4	12.317	152	142194	54.504	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 109.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041036.D
Acq On : 12 Apr 2024 17:29
Operator : JC/MD
Sample : P2049-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0RH7

Quant Time: Apr 13 01:41:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:36:25 2024
Response via : Initial Calibration

