

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041046.D
 Acq On : 12 Apr 2024 21:19
 Operator : JC/MD
 Sample : P2060-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E1D13

Quant Time: Apr 13 04:23:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 04:06:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	242487	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	231833	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	125968	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	89089	51.654	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.300%
7) Chloroethane-d5	1.672	69	73887	52.527	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.060%
11) 1,1-Dichloroethene-d2	2.307	65	36514	49.587	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.180%
21) 2-Butanone-d5	4.459	46	117894	125.486	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	125.490%
24) Chloroform-d	5.056	84	162017	50.345	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.700%
26) 1,2-Dichloroethane-d4	5.952	65	106763	51.002	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.000%
32) Benzene-d6	5.977	84	324907	51.623	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.240%
36) 1,2-Dichloropropane-d6	7.312	67	95805	52.193	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.380%
41) Toluene-d8	8.647	98	300238	48.828	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.660%
43) trans-1,3-Dichloroprop...	8.952	79	43683	47.289	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.580%
47) 2-Hexanone-d5	9.384	63	81996	105.949	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.950%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	144801	53.253	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.500%
66) 1,2-Dichlorobenzene-d4	12.317	152	137797	54.532	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.060%
Target Compounds						
					Qvalue	
13) Acetone	2.380	43	5719	5.616	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041046.D
Acq On : 12 Apr 2024 21:19
Operator : JC/MD
Sample : P2060-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E1D13

Quant Time: Apr 13 04:23:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 04:06:48 2024
Response via : Initial Calibration

