

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041048.D
 Acq On : 12 Apr 2024 22:05
 Operator : JC/MD
 Sample : P2098-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MC0R76

Quant Time: Apr 13 04:23:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 04:06:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	247001	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	235624	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	127510	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	89343	50.854	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.700%
7) Chloroethane-d5	1.672	69	75217	52.496	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.000%
11) 1,1-Dichloroethene-d2	2.306	65	37121	49.490	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.980%
21) 2-Butanone-d5	4.458	46	124600	130.200	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	130.200%#
24) Chloroform-d	5.056	84	164335	50.132	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.260%
26) 1,2-Dichloroethane-d4	5.952	65	109492	51.350	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.700%
32) Benzene-d6	5.976	84	326089	50.977	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.960%
36) 1,2-Dichloropropane-d6	7.312	67	97793	52.418	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.840%
41) Toluene-d8	8.647	98	303867	48.624	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.240%
43) trans-1,3-Dichloroprop...	8.951	79	43312	46.133	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.260%
47) 2-Hexanone-d5	9.384	63	85778	109.053	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.050%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	144754	52.379	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.760%
66) 1,2-Dichlorobenzene-d4	12.317	152	135354	52.917	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041048.D
Acq On : 12 Apr 2024 22:05
Operator : JC/MD
Sample : P2098-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC0R76

Quant Time: Apr 13 04:23:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 04:06:48 2024
Response via : Initial Calibration

