

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041058.D
 Acq On : 13 Apr 2024 01:55
 Operator : JC/MD
 Sample : P2098-10 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MCORA2

Quant Time: Apr 13 04:25:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 04:06:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	245032	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	237183	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	122996	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	89679	51.456	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.920%	
7) Chloroethane-d5	1.666	69	76465	53.796	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 107.600%	
11) 1,1-Dichloroethene-d2	2.306	65	37526	50.432	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 100.860%	
21) 2-Butanone-d5	4.452	46	134235	141.395	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 141.400%#	
24) Chloroform-d	5.056	84	170995	52.583	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.160%	
26) 1,2-Dichloroethane-d4	5.952	65	113023	53.432	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 106.860%	
32) Benzene-d6	5.970	84	340650	52.903	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.800%	
36) 1,2-Dichloropropane-d6	7.306	67	100396	53.460	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 106.920%	
41) Toluene-d8	8.647	98	314507	49.995	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 100.000%	
43) trans-1,3-Dichloroprop...	8.952	79	44485	47.071	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 94.140%	
47) 2-Hexanone-d5	9.384	63	93770	118.430	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 118.430%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	152820	54.934	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 109.860%	
66) 1,2-Dichlorobenzene-d4	12.317	152	139867	56.689	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 113.380%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041058.D
Acq On : 13 Apr 2024 01:55
Operator : JC/MD
Sample : P2098-10 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC0RA2

Quant Time: Apr 13 04:25:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 04:06:48 2024
Response via : Initial Calibration

