

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
 Data File : VX041059.D
 Acq On : 13 Apr 2024 02:18
 Operator : JC/MD
 Sample : P2098-11 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MCORA6

Quant Time: Apr 13 04:25:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 04:06:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	234830	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	223156	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	117764	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	84840	50.794	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.580%
7) Chloroethane-d5	1.672	69	71895	52.778	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.560%
11) 1,1-Dichloroethene-d2	2.307	65	36039	50.538	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.080%
21) 2-Butanone-d5	4.452	46	126848	139.419	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	139.420%#
24) Chloroform-d	5.056	84	160564	51.520	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.040%
26) 1,2-Dichloroethane-d4	5.958	65	106925	52.745	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.480%
32) Benzene-d6	5.977	84	319021	52.658	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.320%
36) 1,2-Dichloropropane-d6	7.306	67	94518	53.494	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.980%
41) Toluene-d8	8.647	98	294796	49.808	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.620%
43) trans-1,3-Dichloroprop...	8.952	79	41914	47.138	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.280%
47) 2-Hexanone-d5	9.384	63	87199	117.053	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	117.050%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	141860	54.200	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.400%
66) 1,2-Dichlorobenzene-d4	12.317	152	133269	56.414	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041224\
Data File : VX041059.D
Acq On : 13 Apr 2024 02:18
Operator : JC/MD
Sample : P2098-11 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MC0RA6

Quant Time: Apr 13 04:25:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 04:06:48 2024
Response via : Initial Calibration

