

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041319\
 Data File : VX008930.D
 Acq On : 13 Apr 2019 14:35
 Operator : JC/SP
 Sample : K2381-02DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-COMP-2DL

Manual Integrations
 APPROVED

MMDadoda
 4/15/2019 12:55:25 PM

Quant Time: Apr 15 02:40:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	176946	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	282103	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	230862	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	105098	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114615	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	
35) Dibromofluoromethane	5.49	113	83275	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
50) Toluene-d8	8.71	98	345275	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
62) 4-Bromofluorobenzene	11.13	95	111955	43.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.92%	
Target Compounds						
					Qvalue	
16) Acetone	2.46	43	284073	191.102	ug/l	100
18) Methyl Acetate	2.78	43	139573	38.714	ug/l	100
25) 2-Butanone	4.69	43	33694	14.341	ug/l	97
37) Ethyl Acetate	4.84	43	150531	36.572	ug/l	98
52) Toluene	8.79	92	2574	0.475	ug/l	94
64) Tetrachloroethene	9.34	164	6457	3.775	ug/l	95
68) m/p-Xylenes	10.36	106	1377	0.401	ug/l	93
78) n-propylbenzene	11.36	91	3358	0.331	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	7356	1.006	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	29081	3.928	ug/l	99
85) sec-Butylbenzene	11.94	105	2543	0.304	ug/l	99
86) p-Isopropyltoluene	12.06	119	2593m	0.346	ug/l	
89) n-Butylbenzene	12.38	91	7139	1.015	ug/l #	47
95) Naphthalene	13.83	128	53538	6.190	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041319\
Data File : VX008930.D
Acq On : 13 Apr 2019 14:35
Operator : JC/SP
Sample : K2381-02DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-COMP-2DL

Manual Integrations
APPROVED

MMDadoda
4/15/2019 12:55:25 PM

Quant Time: Apr 15 02:40:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

