

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041319\
 Data File : VX008941.D
 Acq On : 13 Apr 2019 18:52
 Operator : JC/SP
 Sample : K2385-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-14-190410

Quant Time: Apr 15 04:21:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	196118	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	311662	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	271231	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	117830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126373	44.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.36%	
35) Dibromofluoromethane	5.49	113	92788	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
50) Toluene-d8	8.71	98	388758	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	11.13	95	129644	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	1004	0.409	ug/l	87
27) cis-1,2-Dichloroethene	4.60	96	54705	19.401	ug/l	88
30) Chloroform	5.21	83	1772	0.401	ug/l	96
44) Trichloroethene	7.21	130	31219	13.207	ug/l	95
64) Tetrachloroethene	9.34	164	306739	152.642	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041319\
Data File : VX008941.D
Acq On : 13 Apr 2019 18:52
Operator : JC/SP
Sample : K2385-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M-14-190410

Quant Time: Apr 15 04:21:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

