


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX041319\
 Data File : VX008924.D
 Acq On : 13 Apr 2019 10:03
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Title : SW846 8260
 Last Update : Thu Apr 04 04:47:41 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	10409	PASS
75	95	30	60	50.3	25800	PASS
95	95	100	100	100.0	51256	PASS
96	95	5	9	6.9	3528	PASS
173	174	0.00	2	0.5	201	PASS
174	95	50	100	78.3	40136	PASS
175	174	5	9	7.4	2967	PASS
176	174	95	101	98.0	39314	PASS
177	176	5	9	6.9	2704	PASS