

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041320\
 Data File : VX015647.D
 Acq On : 13 Apr 2020 17:11
 Operator : JC/SP
 Sample : L2241-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200409

Quant Time: Apr 14 01:35:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1094781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1660533	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1602915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	876188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	671324	43.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.96%	
35) Dibromofluoromethane	5.47	113	510974	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
50) Toluene-d8	8.70	98	2000132	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.12	95	832952	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
Target Compounds						
16) Acetone	2.42	43	46912	7.072	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041320\
Data File : VX015647.D
Acq On : 13 Apr 2020 17:11
Operator : JC/SP
Sample : L2241-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-200409

Quant Time: Apr 14 01:35:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

