

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041320\
 Data File : VX015649.D
 Acq On : 13 Apr 2020 17:57
 Operator : JC/SP
 Sample : L2241-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-SB-200409

Quant Time: Apr 14 01:41:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1082929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1701395	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1631900	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	889415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	677512	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	
35) Dibromofluoromethane	5.47	113	513943	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
50) Toluene-d8	8.70	98	2031808	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
62) 4-Bromofluorobenzene	11.12	95	854496	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
Target Compounds						
16) Acetone	2.42	43	20484	3.122	ug/l	96
64) Tetrachloroethene	9.32	164	13341	1.083	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041320\
Data File : VX015649.D
Acq On : 13 Apr 2020 17:57
Operator : JC/SP
Sample : L2241-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-SB-200409

Quant Time: Apr 14 01:41:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

