

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041320\
 Data File : VX015650.D
 Acq On : 13 Apr 2020 18:19
 Operator : JC/SP
 Sample : L2239-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z5-S1GW-200407

Quant Time: Apr 14 01:46:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1054206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1659860	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1593518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	912760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	671473	45.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.32%	
35) Dibromofluoromethane	5.47	113	504969	47.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.18%	
50) Toluene-d8	8.70	98	1989728	50.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.12	95	853141	53.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.48%	
Target Compounds						
16) Acetone	2.42	43	84237	13.187	ug/l	97
25) 2-Butanone	4.64	43	253272	24.654	ug/l	99
29) Tetrahydrofuran	5.10	42	48097	7.283	ug/l	96
44) Trichloroethene	7.20	130	8010	0.656	ug/l	87
47) Bromodichloromethane	7.87	83	12588	0.735	ug/l #	88
60) Dibromochloromethane	9.57	129	20041	1.474	ug/l	97
71) Bromoform	10.84	173	18452	1.597	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041320\
Data File : VX015650.D
Acq On : 13 Apr 2020 18:19
Operator : JC/SP
Sample : L2239-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z5-S1GW-200407

Quant Time: Apr 14 01:46:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

