

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
 Data File : VX028093.D
 Acq On : 13 Apr 2022 16:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Apr 14 04:06:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	186404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	307492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	127356	50.360	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.720%	
35) Dibromofluoromethane	5.391	113	103939	51.585	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.180%	
50) Toluene-d8	8.653	98	392173	52.187	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.380%	
62) 4-Bromofluorobenzene	11.085	95	144069	51.346	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.700%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	89570	49.489	ug/l	100
3) Chloromethane	1.288	50	109641	54.737	ug/l	99
4) Vinyl Chloride	1.374	62	135524	56.278	ug/l	100
5) Bromomethane	1.605	94	154021	60.752	ug/l	97
6) Chloroethane	1.679	64	117916	59.043	ug/l	97
7) Trichlorofluoromethane	1.886	101	326628	69.160	ug/l	97
8) Diethyl Ether	2.136	74	114195	63.095	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.331	101	93487	49.757	ug/l	97
10) Methyl Iodide	2.453	142	109918	50.237	ug/l	97
11) Tert butyl alcohol	2.971	59	144253	245.739	ug/l	100
12) 1,1-Dichloroethene	2.319	96	92137	50.077	ug/l	99
13) Acrolein	2.233	56	86093	277.596	ug/l	100
14) Allyl chloride	2.666	41	155859	47.864	ug/l	88
15) Acrylonitrile	3.063	53	300790	247.279	ug/l	98
16) Acetone	2.380	43	252367	239.518	ug/l	92
17) Carbon Disulfide	2.514	76	237506	49.121	ug/l	100
18) Methyl Acetate	2.703	43	137042	48.743	ug/l	94
19) Methyl tert-butyl Ether	3.117	73	358475	51.878	ug/l	98
20) Methylene Chloride	2.788	84	105616	50.326	ug/l	93
21) trans-1,2-Dichloroethene	3.093	96	107324	48.694	ug/l	97
22) Diisopropyl ether	3.764	45	354336	50.709	ug/l	89
23) Vinyl Acetate	3.727	43	1609080	259.151	ug/l	95
24) 1,1-Dichloroethane	3.611	63	192539	50.900	ug/l	100
25) 2-Butanone	4.556	43	444508	244.066	ug/l	95
26) 2,2-Dichloropropane	4.477	77	170389	48.519	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	125273	50.423	ug/l	98
28) Bromochloromethane	4.904	49	78249	48.125	ug/l	92
29) Tetrahydrofuran	5.007	42	295359	248.120	ug/l	91
30) Chloroform	5.099	83	211178	50.204	ug/l	94
31) Cyclohexane	5.471	56	168518	48.463	ug/l	93
32) 1,1,1-Trichloroethane	5.385	97	188495	49.477	ug/l	98
36) 1,1-Dichloropropene	5.696	75	148691	48.352	ug/l	99
37) Ethyl Acetate	4.715	43	177770	49.624	ug/l	96
38) Carbon Tetrachloride	5.678	117	164840	50.210	ug/l	99
39) Methylcyclohexane	7.385	83	178987	49.238	ug/l	98
40) Benzene	6.044	78	446309	49.911	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
 Data File : VX028093.D
 Acq On : 13 Apr 2022 16:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Apr 14 04:06:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	90834	47.660	ug/l	94
42) 1,2-Dichloroethane	6.092	62	169553	50.214	ug/l	98
43) Isopropyl Acetate	6.342	43	291104	48.776	ug/l	93
44) Trichloroethene	7.129	130	120399	50.633	ug/l	93
45) 1,2-Dichloropropane	7.434	63	111157	49.326	ug/l	100
46) Dibromomethane	7.586	93	82597	49.683	ug/l	95
47) Bromodichloromethane	7.824	83	159473	49.633	ug/l	98
48) Methyl methacrylate	7.696	41	132223	48.531	ug/l #	84
49) 1,4-Dioxane	7.671	88	64608	1062.234	ug/l	93
51) 4-Methyl-2-Pentanone	8.574	43	916219	240.779	ug/l	91
52) Toluene	8.720	92	295230	49.487	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	182981	50.419	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	188243	50.008	ug/l #	85
55) 1,1,2-Trichloroethane	9.153	97	117312	48.507	ug/l	94
56) Ethyl methacrylate	9.116	69	189136	49.753	ug/l #	79
57) 1,3-Dichloropropane	9.311	76	199893	49.858	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.245	63	441463	257.291	ug/l	100
59) 2-Hexanone	9.433	43	717798	240.857	ug/l	88
60) Dibromochloromethane	9.525	129	128127	49.560	ug/l	95
61) 1,2-Dibromoethane	9.610	107	131509	49.928	ug/l	99
64) Tetrachloroethene	9.275	164	112803	49.595	ug/l	98
65) Chlorobenzene	10.080	112	314311	50.250	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	119317	51.627	ug/l	98
67) Ethyl Benzene	10.195	91	572392	51.965	ug/l	98
68) m/p-Xylenes	10.305	106	459241	99.846	ug/l	100
69) o-Xylene	10.640	106	223607	50.013	ug/l	99
70) Styrene	10.659	104	374320	43.155	ug/l	98
71) Bromoform	10.805	173	94767	43.108	ug/l	98
73) Isopropylbenzene	10.964	105	580431	50.023	ug/l	100
74) N-amyl acetate	10.842	43	263061	52.458	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	11.213	83	196535	48.484	ug/l	100
76) 1,2,3-Trichloropropane	11.244	75	226249	62.221	ug/l	90
77) Bromobenzene	11.201	156	144296	50.906	ug/l	94
78) n-propylbenzene	11.305	91	689135	51.103	ug/l	100
79) 2-Chlorotoluene	11.366	91	403676	49.362	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	522823	52.161	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	65688	50.879	ug/l	93
82) 4-Chlorotoluene	11.457	91	489116	50.806	ug/l	98
83) tert-Butylbenzene	11.719	119	500087	51.567	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	510884	52.061	ug/l	99
85) sec-Butylbenzene	11.890	105	621667	50.226	ug/l	99
86) p-Isopropyltoluene	12.012	119	537032	51.205	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	274955	49.782	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	293327	50.628	ug/l	99
89) n-Butylbenzene	12.335	91	464476	51.741	ug/l	98
90) Hexachloroethane	12.543	117	89452	50.763	ug/l	92
91) 1,2-Dichlorobenzene	12.335	146	281333	50.977	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	47021	52.726	ug/l	94
93) 1,2,4-Trichlorobenzene	13.591	180	167580	54.303	ug/l	98
94) Hexachlorobutadiene	13.725	225	63813	51.245	ug/l	92
95) Naphthalene	13.774	128	606879	56.306	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	168514	55.177	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
Data File : VX028093.D
Acq On : 13 Apr 2022 16:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Apr 14 04:06:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
 Data File : VX028093.D
 Acq On : 13 Apr 2022 16:24
 Operator : JC/MD
 Sample : VSTDC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC050EC

Quant Time: Apr 14 04:06:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 186404
Ion Ratio Lower Upper
168 100
99 56.3 41.7 62.5



#2
Dichlorodifluoromethane
Concen: 49.489 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.001 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 85 Resp: 89570
Ion Ratio Lower Upper
85 100
87 32.4 16.2 48.6









#3
Chloromethane
Concen: 54.737 ug/l
RT: 1.288 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 109641
Ion Ratio Lower Upper
50 100
52 32.3 26.4 39.6



#4
Vinyl Chloride
Concen: 56.278 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 62 Resp: 135524
Ion Ratio Lower Upper
62 100
64 31.8 25.5 38.3









#5
Bromomethane
Concen: 60.752 ug/l
RT: 1.605 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 154021
Ion Ratio Lower Upper
94 100
96 96.7 75.3 112.9



#6
Chloroethane
Concen: 59.043 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 64 Resp: 117916
Ion Ratio Lower Upper
64 100
66 33.1 25.3 37.9









#7
 Trichlorofluoromethane
 Concen: 69.160 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 326628
 Ion Ratio Lower Upper
 101 100
 103 65.2 50.5 75.7



#8
 Diethyl Ether
 Concen: 63.095 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24

Tgt Ion: 74 Resp: 114195
 Ion Ratio Lower Upper
 74 100
 45 79.6 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.757 ug/l
RT: 2.331 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 93487
Ion Ratio Lower Upper
101 100
85 46.3 33.5 50.3
151 77.9 61.8 92.6



#10
Methyl Iodide
Concen: 50.237 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:142 Resp: 109918
Ion Ratio Lower Upper
142 100
127 44.5 33.9 50.9
141 14.1 11.4 17.2









#11
Tert butyl alcohol
Concen: 245.739 ug/l
RT: 2.971 min Scan# 308
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 144253
Ion Ratio Lower Upper
59 100
57 10.4 8.3 12.5



#12
1,1-Dichloroethene
Concen: 50.077 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 96 Resp: 92137
Ion Ratio Lower Upper
96 100
61 154.6 124.6 187.0
98 63.4 51.0 76.4





#13

Acrolein

Concen: 277.596 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 56 Resp: 86093

Ion Ratio Lower Upper

56 100

55 69.0 55.0 82.6



#14

Allyl chloride

Concen: 47.864 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion: 41 Resp: 155859


Ion Ratio Lower Upper

41 100

39 75.1 50.4 75.6

76 37.2 32.7 49.1









#15
Acrylonitrile
Concen: 247.279 ug/l
RT: 3.063 min Scan# 31
Delta R.T. 0.001 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 300790
Ion Ratio Lower Upper
53 100
52 83.8 65.7 98.5
51 36.1 28.4 42.6



#16
Acetone
Concen: 239.518 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 43 Resp: 252367
Ion Ratio Lower Upper
43 100
58 31.1 28.8 43.2









#17
Carbon Disulfide
Concen: 49.121 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 237506
Ion Ratio Lower Upper
76 100
78 9.0 7.2 10.8



#18
Methyl Acetate
Concen: 48.743 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 43 Resp: 137042
Ion Ratio Lower Upper
43 100
74 25.2 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 51.878 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 358475
Ion Ratio Lower Upper
73 100
57 20.0 16.8 25.2



#20
Methylene Chloride
Concen: 50.326 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 84 Resp: 105616
Ion Ratio Lower Upper
84 100
49 120.0 85.5 128.3
51 35.4 26.6 39.8
86 64.4 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 48.694 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 107324
Ion Ratio Lower Upper
96 100
61 134.9 105.4 158.2
98 65.1 50.0 75.0



#22
Diisopropyl ether
Concen: 50.709 ug/l
RT: 3.764 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 45 Resp: 354336
Ion Ratio Lower Upper
45 100
43 75.5 50.8 76.2
87 30.2 26.6 39.8
59 11.2 10.5 15.7









#23
 Vinyl Acetate
 Concen: 259.151 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1609080
 Ion Ratio Lower Upper
 43 100
 86 11.7 10.8 16.2



#24
 1,1-Dichloroethane
 Concen: 50.900 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24

Tgt Ion: 63 Resp: 192539
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.9 11.7
 100 5.0 2.4 7.2





#25

2-Butanone

Concen: 244.066 ug/l

RT: 4.556 min Scan# 5

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 444508

Ion Ratio Lower Upper

43 100

72 27.7 24.4 36.6



#26

2,2-Dichloropropane

Concen: 48.519 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 77 Resp: 170389

Ion Ratio Lower Upper

77 100

97 22.7 11.8 35.4









#27
 cis-1,2-Dichloroethene
 Concen: 50.423 ug/l
 RT: 4.495 min Scan# 511
 Delta R.T. 0.006 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 125273
 Ion Ratio Lower Upper
 96 100
 61 138.7 0.0 283.8
 98 65.2 0.0 129.8



#28
 Bromochloromethane
 Concen: 48.125 ug/l
 RT: 4.904 min Scan# 626
 Delta R.T. -0.000 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24

Tgt Ion: 49 Resp: 78249
 Ion Ratio Lower Upper
 49 100
 129 2.4 0.0 4.4
 130 80.8 70.7 106.1





#29

Tetrahydrofuran

Concen: 248.120 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX028093.D


Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 42 Resp: 295359

Ion Ratio Lower Upper

42 100

72 47.6 44.1 66.1

71 44.9 40.5 60.7



#30

Chloroform

Concen: 50.204 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 83 Resp: 211178

Ion Ratio Lower Upper

83 100

85 61.0 52.7 79.1









#31
Cyclohexane
Concen: 48.463 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 168518
Ion Ratio Lower Upper
56 100
69 30.2 26.7 40.1
84 88.7 77.2 115.8



#32
1,1,1-Trichloroethane
Concen: 49.477 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 97 Resp: 188495
Ion Ratio Lower Upper
97 100
99 64.4 51.8 77.6
61 45.2 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 50.360 ug/l

RT: 5.964 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 127356

Ion Ratio Lower Upper

65 100

67 51.0 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.006 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion: 114 Resp: 307492


Ion Ratio Lower Upper

114 100

63 18.6 0.0 36.8

88 15.1 0.0 32.4





#35

Dibromofluoromethane

Concen: 51.585 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 103939

Ion Ratio Lower Upper

113 100

111 104.8 82.8 124.2

192 20.0 16.9 25.3



#36

1,1-Dichloropropene

Concen: 48.352 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion: 75 Resp: 148691


Ion Ratio Lower Upper

75 100

110 37.0 18.3 54.8

77 31.5 24.8 37.2









#37
Ethyl Acetate
Concen: 49.624 ug/l
RT: 4.715 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 177770
Ion Ratio Lower Upper
43 100
61 14.3 12.6 19.0
70 10.8 10.2 15.2



#38
Carbon Tetrachloride
Concen: 50.210 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 117 Resp: 164840
Ion Ratio Lower Upper
117 100
119 95.1 77.3 115.9
121 29.4 24.2 36.2





#39

Methylcyclohexane

Concen: 49.238 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 178987

Ion Ratio Lower Upper

83 100

55 68.1 54.6 81.8

98 49.5 37.4 56.2



#40

Benzene

Concen: 49.911 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.007 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 78 Resp: 446309

Ion Ratio Lower Upper

78 100

77 23.6 19.3 28.9









#41
Methacrylonitrile
Concen: 47.660 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 41 Resp: 90834
Ion Ratio Lower Upper
41 100
39 55.9 43.0 64.4
67 78.1 70.2 105.4
52 33.3 26.0 39.0



#42
1,2-Dichloroethane
Concen: 50.214 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 62 Resp: 169553
Ion Ratio Lower Upper
62 100
98 9.4 0.0 20.4









#43
Isopropyl Acetate
Concen: 48.776 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 291104
Ion Ratio Lower Upper
43 100
61 19.1 18.5 27.7
87 14.2 13.0 19.6



#44
Trichloroethene
Concen: 50.633 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 130 Resp: 120399
Ion Ratio Lower Upper
130 100
95 101.3 0.0 188.8





#45

1,2-Dichloropropane

Concen: 49.326 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 111157

Ion Ratio Lower Upper

63 100

65 32.0 25.8 38.6



#46

Dibromomethane

Concen: 49.683 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion: 93 Resp: 82597


Ion Ratio Lower Upper

93 100

95 83.4 65.8 98.6

174 98.5 85.4 128.2









#47
Bromodichloromethane
Concen: 49.633 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 159473
Ion Ratio Lower Upper
83 100
85 63.9 52.2 78.2
127 9.1 6.8 10.2



#48
Methyl methacrylate
Concen: 48.531 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 41 Resp: 132223
Ion Ratio Lower Upper
41 100
69 86.5 82.5 123.7
39 66.6 43.8 65.8#





#49

1,4-Dioxane

Concen: 1062.234 ug/l

RT: 7.671 min Scan# 1080

Delta R.T. 0.012 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 64608

Ion Ratio Lower Upper

88 100

43 31.3 21.9 32.9

58 61.9 53.7 80.5



#50

Toluene-d8

Concen: 52.187 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 98 Resp: 392173

Ion Ratio Lower Upper

98 100

100 65.6 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 240.779 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 916219
Ion Ratio Lower Upper
43 100
58 39.5 36.1 54.1



#52
Toluene
Concen: 49.487 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 92 Resp: 295230
Ion Ratio Lower Upper
92 100
91 171.7 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 50.419 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 182981

Ion Ratio Lower Upper

75 100

77 31.6 25.4 38.0



#54

cis-1,3-Dichloropropene

Concen: 50.008 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion: 75 Resp: 188243


Ion Ratio Lower Upper

75 100

77 32.2 24.8 37.2

39 52.7 29.6 44.4#





#55

1,1,2-Trichloroethane

Concen: 48.507 ug/l

RT: 9.153 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 117312





Ion Ratio Lower Upper

97 100

83 90.2 66.2 99.4

85 55.6 42.9 64.3

99 64.9 49.0 73.6



#56

Ethyl methacrylate

Concen: 49.753 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion: 69 Resp: 189136


Ion Ratio Lower Upper

69 100

41 65.5 43.8 65.6

39 44.7 22.2 33.2#





#57

1,3-Dichloropropane

Concen: 49.858 ug/l

RT: 9.311 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 199893

Ion Ratio Lower Upper

76 100

78 32.1 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 257.291 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. 0.001 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 63 Resp: 441463

Ion Ratio Lower Upper

63 100

106 29.8 23.9 35.9





#59

2-Hexanone

Concen: 240.857 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 717798

Ion Ratio Lower Upper

43 100

58 54.6 31.9 95.7



#60

Dibromochloromethane

Concen: 49.560 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 129 Resp: 128127

Ion Ratio Lower Upper

129 100

127 76.1 40.3 120.8









#61
1,2-Dibromoethane
Concen: 49.928 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 131509
Ion Ratio Lower Upper
107 100
109 92.4 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 51.346 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 95 Resp: 144069
Ion Ratio Lower Upper
95 100
174 83.3 0.0 163.0
176 80.9 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 286670
Ion Ratio Lower Upper
117 100
82 55.4 41.8 62.8
119 31.9 24.8 37.2



#64
Tetrachloroethene
Concen: 49.595 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:164 Resp: 112803
Ion Ratio Lower Upper
164 100
166 127.1 101.2 151.8
129 96.5 73.0 109.4
131 92.2 74.2 111.2









#65
Chlorobenzene
Concen: 50.250 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.001 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 314311
Ion Ratio Lower Upper
112 100
114 31.9 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 51.627 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:131 Resp: 119317
Ion Ratio Lower Upper
131 100
133 94.0 48.3 144.8
119 64.9 32.6 98.0









#67
Ethyl Benzene
Concen: 51.965 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 572392
Ion Ratio Lower Upper
91 100
106 31.8 24.6 36.8



#68
m/p-Xylenes
Concen: 99.846 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:106 Resp: 459241
Ion Ratio Lower Upper
106 100
91 200.1 160.4 240.6









#69
o-Xylene
Concen: 50.013 ug/l
RT: 10.640 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 223607
Ion Ratio Lower Upper
106 100
91 210.5 106.1 318.1



#70
Styrene
Concen: 43.155 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:104 Resp: 374320
Ion Ratio Lower Upper
104 100
78 49.8 38.6 57.8
103 55.9 43.8 65.8









#71
Bromoform
Concen: 43.108 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.006 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 94767
Ion Ratio Lower Upper
173 100
175 49.6 24.1 72.2
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:152 Resp: 163335
Ion Ratio Lower Upper
152 100
115 83.9 37.5 112.7
150 177.3 0.0 338.2





#73

Isopropylbenzene

Concen: 50.023 ug/l

RT: 10.964 min Scan# 1

Delta R.T. 0.001 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 580431

Ion Ratio Lower Upper

105 100

120 27.1 13.5 40.5



#74

N-ethyl acetate

Concen: 52.458 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Tgt Ion: 43 Resp: 263061




Ion Ratio Lower Upper


43 100

70 43.2 42.5 63.7

55 23.0 23.1 34.7#

61 24.2 22.8 34.2









#75
1,1,2,2-Tetrachloroethane
Concen: 48.484 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 196535
Ion Ratio Lower Upper
83 100
131 10.4 5.1 15.3
85 64.9 32.6 97.7



#76
1,2,3-Trichloropropane
Concen: 62.221 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 75 Resp: 226249
Ion Ratio Lower Upper
75 100
77 32.8 19.4 58.2





#77

Bromobenzene

Concen: 50.906 ug/l

RT: 11.201 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 144296

Ion Ratio Lower Upper

156 100

77 148.7 78.5 235.3

158 94.2 50.2 150.7



#78

n-propylbenzene

Concen: 51.103 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 91 Resp: 689135

Ion Ratio Lower Upper

91 100

120 23.0 11.6 34.8









#79
2-Chlorotoluene
Concen: 49.362 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 403676
Ion Ratio Lower Upper
91 100
126 35.0 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 52.161 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 105 Resp: 522823
Ion Ratio Lower Upper
105 100
120 48.7 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 50.879 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 65688
Ion Ratio Lower Upper
75 100
53 96.3 69.0 103.4
89 45.7 37.0 55.6



#82
4-Chlorotoluene
Concen: 50.806 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 91 Resp: 489116
Ion Ratio Lower Upper
91 100
126 30.8 14.9 44.5









#83
 tert-Butylbenzene
 Concen: 51.567 ug/l
 RT: 11.719 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:119 Resp: 500087
 Ion Ratio Lower Upper
 119 100
 91 56.7 26.9 80.6
 134 23.0 11.4 34.2



#84
 1,2,4-Trimethylbenzene
 Concen: 52.061 ug/l
 RT: 11.756 min Scan# 1750
 Delta R.T. 0.000 min
 Lab File: VX028093.D
 Acq: 13 Apr 2022 16:24

Tgt Ion:105 Resp: 510884
 Ion Ratio Lower Upper
 105 100
 120 44.6 22.6 67.8





#85

sec-Butylbenzene

Concen: 50.226 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:105 Resp: 621667

Ion Ratio Lower Upper

105 100

134 20.8 10.2 30.4



#86

p-Isopropyltoluene

Concen: 51.205 ug/l

RT: 12.012 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion:119 Resp: 537032


Ion Ratio Lower Upper

119 100

134 27.3 13.0 39.0

91 24.6 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 49.782 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 274955

Ion Ratio Lower Upper

146 100

111 40.3 19.7 59.1

148 64.8 31.6 94.7



#88

1,4-Dichlorobenzene

Concen: 50.628 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24




Tgt Ion:146 Resp: 293327


Ion Ratio Lower Upper

146 100

111 39.1 19.2 57.6

148 63.4 32.3 96.9





#89

n-Butylbenzene

Concen: 51.741 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 91 Resp: 464476

Ion Ratio Lower Upper

91 100

92 53.4 27.8 83.3

134 26.1 13.3 39.8



#90

Hexachloroethane

Concen: 50.763 ug/l

RT: 12.543 min Scan# 1879

Delta R.T. 0.001 min

Lab File: VX028093.D




Acq: 13 Apr 2022 16:24


Tgt Ion: 117 Resp: 89452

Ion Ratio Lower Upper

117 100

201 71.9 39.5 118.5









#91
1,2-Dichlorobenzene
Concen: 50.977 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 281333
Ion Ratio Lower Upper
146 100
111 40.5 21.0 63.0
148 64.3 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.726 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion: 75 Resp: 47021
Ion Ratio Lower Upper
75 100
155 87.1 47.0 141.2
157 110.8 57.9 173.6









#93
1,2,4-Trichlorobenzene
Concen: 54.303 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:180 Resp: 167580
Ion Ratio Lower Upper
180 100
182 93.7 47.7 143.1
145 29.8 15.2 45.5



#94
Hexachlorobutadiene
Concen: 51.245 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX028093.D
Acq: 13 Apr 2022 16:24

Tgt Ion:225 Resp: 63813
Ion Ratio Lower Upper
225 100
223 60.9 34.2 102.5
227 62.8 33.9 101.7





#95

Naphthalene

Concen: 56.306 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 606879

Ion Ratio Lower Upper

128 100

127 13.1 10.2 15.4

129 10.9 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 55.177 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX028093.D

Acq: 13 Apr 2022 16:24

Tgt Ion:180 Resp: 168514

Ion Ratio Lower Upper

180 100

182 94.6 50.0 149.8

145 31.7 16.2 48.5

