

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
 Data File : VX028087.D
 Acq On : 13 Apr 2022 14:04
 Operator : JC/MD
 Sample : N2328-03ME
 Misc : 6.15g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-3ME

Quant Time: Apr 14 05:03:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	247563	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	415170	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	198803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158655	47.238	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.480%
35) Dibromofluoromethane	5.385	113	135083	49.654	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.300%
50) Toluene-d8	8.647	98	510405	50.305	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.085	95	171021	45.144	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.280%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.373	83	22598	4.604	ug/l	99
73) Isopropylbenzene	10.964	105	112432	7.961	ug/l	100
78) n-propylbenzene	11.305	91	192984	11.758	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	125517	10.509	ug/l	99
85) sec-Butylbenzene	11.890	105	218531	14.506	ug/l	90
86) p-Isopropyltoluene	12.012	119	24100	1.888	ug/l	93
89) n-Butylbenzene	12.335	91	192997	17.664	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
Data File : VX028087.D
Acq On : 13 Apr 2022 14:04
Operator : JC/MD
Sample : N2328-03ME
Misc : 6.15g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-3ME

Quant Time: Apr 14 05:03:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

