

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
 Data File : VX028079.D
 Acq On : 13 Apr 2022 09:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 14 04:02:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.485	0.476	1.9	98	0.00
3 P	Chloromethane	0.537	0.502	6.5	96	0.00
4 C	Vinyl Chloride	0.646	0.637	1.4#	99	0.00
5 T	Bromomethane	0.724	0.816	-12.7	109	0.00
6 T	Chloroethane	0.552	0.551	0.2	111	0.00
7 T	Trichlorofluoromethane	1.267	1.347	-6.3	108	0.00
8 T	Diethyl Ether	0.485	0.517	-6.6	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.525	-4.2	108	0.00
10 T	Methyl Iodide	0.587	0.538	8.3	85	0.00
11 T	Tert butyl alcohol	0.157	0.143	8.9	92	0.00
12 CM	1,1-Dichloroethene	0.494	0.490	0.8#	105	0.00
13 T	Acrolein	0.091	0.094	-3.3	108	0.00
14 T	Allyl chloride	0.873	0.832	4.7	100	0.00
15 T	Acrylonitrile	0.326	0.321	1.5	96	0.00
16 T	Acetone	0.283	0.315	-11.3	115	0.00
17 T	Carbon Disulfide	1.297	1.354	-4.4	101	0.00
18 T	Methyl Acetate	0.754	0.718	4.8	98	0.00
19 T	Methyl tert-butyl Ether	1.853	1.893	-2.2	100	0.00
20 T	Methylene Chloride	0.563	0.566	-0.5	102	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.583	1.4	101	0.00
22 T	Diisopropyl ether	1.874	1.912	-2.0	99	0.00
23 T	Vinyl Acetate	1.665	1.735	-4.2	98	0.00
24 P	1,1-Dichloroethane	1.015	1.046	-3.1	101	0.00
25 T	2-Butanone	0.489	0.492	-0.6	99	0.00
26 T	2,2-Dichloropropane	0.942	0.988	-4.9	103	0.00
27 T	cis-1,2-Dichloroethene	0.666	0.672	-0.9	102	0.00
28 T	Bromochloromethane	0.436	0.417	4.4	90	0.00
29 T	Tetrahydrofuran	0.319	0.309	3.1	94	0.00
30 C	Chloroform	1.128	1.143	-1.3#	99	0.00
31 T	Cyclohexane	0.933	0.946	-1.4	99	-0.01
32 T	1,1,1-Trichloroethane	1.022	1.027	-0.5	99	0.00
33 S	1,2-Dichloroethane-d4	0.678	0.624	8.0	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.328	0.320	2.4	93	-0.01
36 T	1,1-Dichloropropene	0.500	0.516	-3.2	100	0.00
37 T	Ethyl Acetate	0.583	0.590	-1.2	97	0.00
38 T	Carbon Tetrachloride	0.534	0.574	-7.5	103	0.00
39 T	Methylcyclohexane	0.591	0.639	-8.1	103	0.00
40 TM	Benzene	1.454	1.523	-4.7	99	0.00
41 T	Methacrylonitrile	0.310	0.295	4.8	94	0.00
42 TM	1,2-Dichloroethane	0.549	0.574	-4.6	99	0.00
43 T	Isopropyl Acetate	0.970	0.973	-0.3	97	0.00
44 TM	Trichloroethene	0.387	0.417	-7.8	102	0.00
45 C	1,2-Dichloropropane	0.366	0.378	-3.3#	99	0.00
46 T	Dibromomethane	0.270	0.278	-3.0	98	0.00
47 T	Bromodichloromethane	0.522	0.555	-6.3	99	0.00
48 T	Methyl methacrylate	0.443	0.436	1.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
 Data File : VX028079.D
 Acq On : 13 Apr 2022 09:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 14 04:02:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	88	0.00
50 S	Toluene-d8	1.222	1.232	-0.8	93	0.00
51 T	4-Methyl-2-Pentanone	0.619	0.623	-0.6	96	0.00
52 CM	Toluene	0.970	1.025	-5.7#	100	0.00
53 T	t-1,3-Dichloropropene	0.590	0.644	-9.2	100	0.00
54 T	cis-1,3-Dichloropropene	0.612	0.647	-5.7	99	0.00
55 T	1,1,2-Trichloroethane	0.393	0.415	-5.6	101	0.00
56 T	Ethyl methacrylate	0.618	0.637	-3.1	97	0.00
57 T	1,3-Dichloropropane	0.652	0.665	-2.0	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.279	0.320	-14.7	105	0.00
59 T	2-Hexanone	0.485	0.501	-3.3	95	0.00
60 T	Dibromochloromethane	0.420	0.457	-8.8	101	0.00
61 T	1,2-Dibromoethane	0.428	0.454	-6.1	102	0.00
62 S	4-Bromofluorobenzene	0.456	0.477	-4.6	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.397	0.425	-7.1	103	0.00
65 PM	Chlorobenzene	1.091	1.145	-4.9	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.403	0.431	-6.9	102	0.00
67 C	Ethyl Benzene	1.921	2.098	-9.2#	102	0.00
68 T	m/p-Xylenes	0.802	0.841	-4.9	101	0.00
69 T	o-Xylene	0.780	0.817	-4.7	101	0.00
70 T	Styrene	1.343	1.384	-3.1	101	0.00
71 P	Bromoform	0.335	0.372	-11.0	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.552	3.734	-5.1	101	0.00
74 T	N-amyl acetate	1.535	1.648	-7.4	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.262	-1.7	100	0.00
76 T	1,2,3-Trichloropropane	1.113	1.169	-5.0	99	0.00
77 T	Bromobenzene	0.868	0.905	-4.3	101	0.00
78 T	n-propylbenzene	4.128	4.434	-7.4	102	0.00
79 T	2-Chlorotoluene	2.503	2.562	-2.4	100	0.00
80 T	1,3,5-Trimethylbenzene	3.068	3.264	-6.4	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.395	0.437	-10.6	103	0.00
82 T	4-Chlorotoluene	2.947	3.080	-4.5	100	0.00
83 T	tert-Butylbenzene	2.969	3.167	-6.7	102	0.00
84 T	1,2,4-Trimethylbenzene	3.004	3.226	-7.4	101	0.00
85 T	sec-Butylbenzene	3.789	4.039	-6.6	101	0.00
86 T	p-Isopropyltoluene	3.211	3.507	-9.2	103	0.00
87 T	1,3-Dichlorobenzene	1.691	1.813	-7.2	103	0.00
88 T	1,4-Dichlorobenzene	1.774	1.830	-3.2	102	0.00
89 T	n-Butylbenzene	2.748	3.093	-12.6	105	0.00
90 T	Hexachloroethane	0.539	0.612	-13.5	103	0.00
91 T	1,2-Dichlorobenzene	1.689	1.775	-5.1	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.273	0.284	-4.0	99	0.00
93 T	1,2,4-Trichlorobenzene	0.945	1.062	-12.4	105	0.00
94 T	Hexachlorobutadiene	0.381	0.418	-9.7	106	0.00
95 T	Naphthalene	3.299	3.671	-11.3	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041322\
Data File : VX028079.D
Acq On : 13 Apr 2022 09:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 14 04:02:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.935	1.047	-12.0	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6